

*TB 9-6625-2357-40

DEPARTMENT OF THE ARMY TECHNICAL BULLETIN

CALIBRATION PROCEDURE FOR DIAL INDICATOR CALIBRATORS FEDERAL PRODUCTS CORP., MODEL 400B-1, THE TIMSCO CO., MODEL TY-1 AND STARRETT CO., MODEL 716 (13589315)

Headquarters, Department of the Army, Washington, DC
14 November 2007

Distribution Statement A: Approved for public release; distribution is unlimited.

REPORTING OF ERRORS AND RECOMMENDING IMPROVEMENTS

You can improve this manual. If you find any mistakes or if you know of a way to improve these procedures, please let us know. Mail your letter or DA Form 2028 (Recommended Changes to Publications and Blank Forms) directly to: Commander, U.S. Army Aviation and Missile Command, ATTN: AMSAM-MMC-MA-NP, Redstone Arsenal, AL 35898-5000. A reply will be furnished to you. You may also send in your comments electronically to our E-mail address: 2028@redstone.army.mil or by fax 256-842-6546/DSN 788-6546. For the World Wide Web use: <https://amcom2028.redstone.army.mil>. Instructions for sending an electronic 2028 can be found at the back of this manual.

SECTION		Paragraph	Page
I.	IDENTIFICATION AND DESCRIPTION		
	Test instrument identification	1	2
	Forms, records, and reports	2	2
	Calibration description.....	3	2
II.	EQUIPMENT REQUIREMENTS		
	Equipment required.....	4	2
	Accessories required	5	2
III.	CALIBRATION PROCESS		
	Preliminary instructions	6	3
	Equipment setup.....	7	3
	Dial indicator calibrator (DIC)		
	linearity accuracy.....	8	3
	Final procedure	9	4

*This bulletin supersedes TB 9-6625-2357-50, dated 17 May 2005.

**SECTION I
IDENTIFICATION AND DESCRIPTION**

1. Test Instrument Identification. This bulletin provides instructions for the calibration of Dial Indicator Calibrators, Federal Products Corp., Model 400B-1, The TIMSCO Co., Model TY-1, and Starrett Co., Model 716 (13589315). DA form 3758-R (Calibration and Repair Requirements Worksheet) was used as the prime data source in compiling these instructions. The equipment being calibrated will be referred to as the TI (test instrument) throughout this bulletin.

a. Model Variations. None

b. Time and Technique. The time required for this calibration is approximately 1 hour, using the physical dimensional technique.

2. Forms, Records, and Reports. Forms, records, and reports required for calibration personnel at all levels are prescribed by TB 750-25.

3. Calibration Description. TI parameters and performance specifications which pertain to this calibration are listed in table 1.

Table 1. Calibration Description

Test instrument parameters	Performance specifications
Range	0.0 to 1.0 in.
Accuracy	± 0.0001 in. over entire range
Resolution	± 0.00005 in.

**SECTION II
EQUIPMENT REQUIREMENTS**

4. Equipment Required. Table 2 identifies the specific equipment to be used in this calibration procedure. This equipment is issued with Secondary Reference Calibration Standards Set, NSN 4931-00-621-7878. Alternate items may be used by the calibrating activity. The items selected must be verified to perform satisfactorily prior to use and must bear evidence of current calibration. The equipment must meet or exceed the minimum use specifications listed in table 2. The accuracies listed in table 2 provide a four-to-one ratio between the standard and TI. Where the four-to-one ratio cannot be met, the actual accuracy of the equipment is shown in parenthesis.

5. Accessories Required. The accessories required for the calibration are common usage accessories, issued as indicated in paragraph 4 above, and are not listed in this calibration procedure. The following peculiar accessories are also required for this calibration: ND281B Display Unit and Shaft diameter adaptor; Split Bushing, Starrett (EDP56007).

Table 2. Minimum Specifications of Equipment Required

Common name	Minimum use specifications	Manufacturer and model (part number)
ELECTRONIC LINEAR TRANSDUCER	Range: 0.0 to 1.0 in. Accuracy: ± 0.000025 in.	Heidenhain, Model MT2501

**SECTION III
CALIBRATION PROCESS**

6. Preliminary Instructions

a. The instructions outlined in paragraphs 6 and 7 are preparatory to the calibration process. Personnel should become familiar with the entire bulletin before beginning the calibration.

b. Items of equipment used in this procedure are referenced within the text by common name as listed in table 2.

c. Unless otherwise specified, verify the result of each test and, whenever the test requirement is not met, take corrective action before continuing with the calibration.

d. Unless otherwise specified, all controls and control settings refer to the TI.

e. Carefully inspect TI anvil for burrs, protrusions, and other defects. If necessary replace TI micrometer head.

7. Equipment Setup

a. Clean TI anvil and electronic linear transducer probe tip with alcohol.

b. Mount electronic linear transducer in TI dial indicator mount using shaft diameter adapter.

c. Connect the electronic linear transducer to the ND281B display unit and connect display unit to 110 V ac.

d. Set electronic linear transducer display unit for a 0.000001 resolution.

NOTE

Zero electronic linear transducer display unit. Ensure the electronic linear transducer and TI has full 1 in. travel.

e. Allow 1 hour equipment stabilization time.

8. Dial Indicator Calibrator (DIC) Linearity Accuracy

a. Performance Check

(1) Set TI to zero position (fully down).

(2) Zero electronic linear transducer display unit.

(3) Adjust TI clockwise (CW) to first TI setting in table 3.

(4) Verify that the electronic linear transducer display unit displays within limits in table 3.

(5) Repeat technique of (3) and (4) for all TI settings listed in table 3.

TB 9-6625-2357-40

Table 3. Calibration Data Chart for Dial Indication Calibrator (DIC)

Rotation Direction	TI Setting	Electronic Linear Transducer Display Unit indication	
		Minimum	Maximum
CW	0.05	0.0499	0.0501
CW	0.1	0.0999	0.1001
CW	0.2	0.1999	0.2001
CW	0.3	0.2999	0.3001
CW	0.4	0.3999	0.4001
CW	0.5	0.4999	0.5001
CW	0.6	0.5999	0.6001
CW	0.7	0.6999	0.7001
CW	0.8	0.7999	0.8001
CW	0.9	0.8999	0.9001
CW	0.95	0.9499	0.9501
CCW	0.5	0.4999	0.5001
CCW	0.0	-0.0001	0.0001

b. Adjustments. No adjustments can be made.

9. Final Procedure

a. Deenergize and disconnect all equipment.

b. Annotate and affix DA label/form in accordance with TB 750-25.

By Order of the Secretary of the Army:

Official:



JOYCE E. MORROW
*Administrative Assistant to the
Secretary of the Army*

0725406

GEORGE W. CASEY, JR.
*General, United States Army
Chief of Staff*

Distribution:

To be distributed in accordance with STD IDS No. RLC-1500, 2 January 2003, requirements for calibration procedure TB 9-6625-2357-40.

Instructions for Submitting an Electronic 2028

The following format must be used if submitting an electronic 2028. The subject line must be exactly the same and all fields must be included; however, only the following fields are mandatory: 1, 3, 4, 5, 6, 7, 8, 9, 10, 13, 15, 16, 17, and 27.

From: "Whomever" whomever@redstone.army.mil
To: <2028@redstone.army.mil

Subject: DA Form 2028

1. **From:** Joe Smith
2. **Unit:** home
3. **Address:** 4300 Park
4. **City:** Hometown
5. **St:** MO
6. **Zip:** 77777
7. **Date Sent:** 19-OCT -93
8. **Pub no:** 55-2840-229-23
9. **Pub Title:** TM
10. **Publication Date:** 04-JUL-85
11. **Change Number:** 7
12. **Submitter Rank:** MSG
13. **Submitter FName:** Joe
14. **Submitter MName:** T
15. **Submitter LName:** Smith
16. **Submitter Phone:** 123-123-1234
17. **Problem:** 1
18. **Page:** 2
19. **Paragraph:** 3
20. **Line:** 4
21. **NSN:** 5
22. **Reference:** 6
23. **Figure:** 7
24. **Table:** 8
25. **Item:** 9
26. **Total:** 123
27. **Text**

This is the text for the problem below line 27.

This fine document...

Was brought to you by me:



[Liberated Manuals -- free army and government manuals](#)

Why do I do it? I am tired of sleazy CD-ROM sellers, who take publicly available information, slap “watermarks” and other junk on it, and sell it. Those masters of search engine manipulation make sure that their sites that sell free information, come up first in search engines. They did not create it... They did not even scan it... Why should they get your money? Why are not letting you give those free manuals to your friends?

I am setting this document FREE. This document was made by the US Government and is NOT protected by Copyright. Feel free to share, republish, sell and so on.

I am not asking you for donations, fees or handouts. If you can, please provide a link to liberatedmanuals.com, so that free manuals come up first in search engines:

<A HREF=<http://www.liberatedmanuals.com/>>Free Military and Government Manuals

- Sincerely
Igor Chudov
<http://igor.chudov.com/>
- [Chicago Machinery Movers](#)