TECHNICAL MANUAL

UNIT

MAINTENANCE MANUAL

FOR

TESTER, DENSITY AND MOISTURE

(SOIL AND ASPHALT)

NUCLEAR METHOD

CAMPBELL PACIFIC MODEL MC-1 (CCE)


(NSN 6635-01-030-6896)

*This manual supersedes TM 5-6635-386-12&P, 10 September 1985.

Approved for public release; distribution is unlimited.

HEADQUARTERS, DEPARTMENT OF THE ARMY

19 AUGUST 1991



WARNING

RADIOACTIVE MATERIAL

Cesium (Cs) 137 and Americium (Am) 241

<u>WARNING</u>: Always wear your beta-gamma or TLD badge when operating the tester and upon entering tester storage areas, and at other times as directed by the Local Radiation Protection Officer (LRPO).

<u>WARNING</u>: Pregnant personnel are not authorized to handle, operate, or maintain the tester, nor serve as the Local Radiation Protection Officer (LRPO).

<u>WARNING</u>: The tester shall immediately be withdrawn from use when the wipe test indicates a reading of 0.1 mrem/hr or more, except in the case of the state of Colorado in which wipe test readings at or above twice the background (0.2 mrem/hr) or when the analysis laboratory test results reveal the presence of 0.005 or more microcuries.

<u>WARNING:</u> A tester is contaminated when a reading of 0.1 mrem/hr or more, except in the case of the state of Colorado in which readings of 0.2 mrem/hr or more, is obtained from the swab. Follow procedure provided in Para 4-31. The swab is also contaminated.

<u>WARNING:</u> The tester contains two sealed sources: Cesium (Cs) 137 and Americium (Am) 241. The sources are classified as special form. The sources are also double-encapsulated in stainless steel. Ten milcures of Cs 137 is in the probe tip and produces gamma radiation. A 50 millicure mixture of Am 241 and Beryllium (Be) is located within the tester base next to the probe and produces fast neutrons.

<u>WARNING</u>: The tester shall only be used by trained operators who are under the direction of a radiological protection officer.

WARNING: The tester operator shall not stand unnecessarily close to the tester.

WARNING: The tester shall be carried by its carrying handle in the safe position.

WARNING: Good radiation protection begins with always concentrating on maintaining maximum distance from the sources, and always operating the tester with speed and efficiency.

WARNING: The tester presents minimum hazard to the operator and the general public when it is used in accordance with procedures provided in this manual.

WARNING: Do not attempt tester repairs. A faulty tester must be returned to a designated manufacturer's repair facility.

<u>WARNING</u>: Insure the tester handle is placed in the safe position prior to placing the tester into the approved carrying case.

<u>WARNING</u>: Dry cleaning solvent (P-D-680), used to clean parts is toxic and flammable. The flashpoint for Type #1 drycleaning solvent is 1 00°F (380C) and for Type #2 is 1380F (50°). Wear protective goggles and gloves and use only in a well ventilated area. Avoid contact with skin, eyes, and clothes. Do not breathe vapors. Do not use near open flame or excessive heat. Do not smoke when using solvent. Failure to do so could cause SERIOUS INJURY. If you become dizzy while using cleaning solvent, get fresh air immediately, and get medical attention. If contact with skin or clothes is made, flush thoroughly with water. If the solvent contacts your eyes--flush immediately with large quantities of water. Medical attention should be received as quickly as possible.

WARNING: While handling the tester, smoking, eating or drinking is not authorized. The same warning applies to tester storage areas where smoking, eating or drinking is not authorized.

<u>WARNING</u>: Under no circumstances shall the testers be conveyed to a property disposal officer. All disposal actions shall be conducted in accordance with AR 385-11.

<u>WARNING</u>: Before proceeding to the next decontamination level, contact the Commander, TACOM, ATTN: AMSTA-CZ for guidance concerning respiratory protection. Failure to wear respiratory protection may result in overexposure to radiation.

Technical Manual

No. 5-6635-386-12&P

HEADQUARTERS
DEPARTMENT OF THE ARMY
Washington D.C., 19 August 1991

UNIT MAINTENANCE MANUAL

FOR

TESTER, DENSITY AND MOISTURE

(SOIL AND ASPHALT) NUCLEAR METHOD CAMPBELL PACIFIC MODEL MC-1 (CCE)

(NSN 6635-01-030-6896)

REPORTING ERRORS AND RECOMMENDING IMPROVEMENTS

You can help improve this manual. If you find any mistakes, or if you know of a way to improve the procedures, please let us know. Mail your letter, DA Form 2028 (Recommended changes to Publications and Blank Forms), or DA Form 2028-2, located in the back of this manual direct to: Commander, US Army Tank-Automotive Command, ATTN: AMSTA-MB, Warren, MI 48937-5000. A reply will be sent to you.

Thin technical manual is an authentication of the manufacturers commercial literature and does not conform with the format and content specified in AR 25-30, Military Publications. This technical manual does. however contain available information that is essential to the operation and maintenance of the equipment.

Portions of this technical manual are copyrighted by Campbell Pacific Nuclear Corporation. 1980. This material may be reproduced by or for the U.S. Government pursuant to the copyright license under DAR 7-1904.9 (a).

NOTE

Standard International (SI) Conversion Chart applicable to units of Radiation Measurement is found in Appendix K.

TABLE OF CONTENTS

		PAGE
SECTION I	GENERAL OPERATING INSTRUCTION	1-1
SECTION II	RADIATION INFORMATION	2-1
SECTION III	OPERATOR MAINTENANCE PROCEDURES	3-1
SECTION IV	SAFETY PROCEDURES	4-1

^{*}This manual supersedes TM 5-6635-386-1 2&P, dated 10 September 1985.

TABLE OF CONTENTS - Continued

		PAGE
SECTION V	SUPPLEMENTAL OPERATING, MAINTENANCE, AND REPAIR	
	PARTS INSTRUCTIONS (SOMARPI)	5-1
	OPERATOR PREVENTIVE MAINTENANCE CHECKS AND	
	SERVICES (PMCS)	5-5
	UNIT PREVENTIVE MAINTENANCE CHECKS AND	
	SERVICES (PMCS)	
APPENDIX A	REFERENCES	
APPENDIX B	MAINTENANCE ALLOCATION CHART	
APPENDIX C	COMPONENT OF END ITEM AND BASIC ISSUE ITEMS LIST	
APPENDIX D	PRESCRIBED LOAD LIST/AUTHORIZED STOCKAGE LIST	
APPENDIX E	EXPENDABLE/DURABLE SUPPLIES AND MATERIALS LIST	
APPENDIX F	MC-1 MAJOR SUB-ASSEMBLES PARTS	F-1
APPENDIX G	LETTER OF INSTRUCTION	
APPENDIX H	SAMPLE FORMAT-MILSTRIP REQUISITION FOR CCE (NON-NSN)	
APPENDIX I	SAMPLE FORMAT-MILSTRIP REQUISITION FOR CCE (NSN)	I-1
APPENDIX J	AREA CONCEPT TESTING	J-1
APPENDIX K	STANDARD INTERNATIONAL CONVERSION CHART	K-1
APPENDIX L	MOISTURE & DENSITY TESTER FIELD DATA WORKSHEET	
	(DA FORM 5448-R, AUG 85)	L-1
APPENDIX M	MOISTURE & DENSITY TESTER (NUCLEAR METHOD)	M-1
APPENDIX N	MOISTURE & DENSITY FIELD TEST SHEET	
	(DA FORM 5450-R AUG 85)	N-1
INDEX		INDEX 1

SECTION I

GENERAL OPERATING INSTRUCTIONS

GENERAL INFORMATION

The Density and Moisture Tester (Soil and Asphalt) measures density and moisture of construction materials using radioactive sources and internal electronics.

Measurements can include the determination of percent compaction of soils and asphalt pavements, density and moisture of concrete, roofing materials, wallboard, woodchips and other construction related materials.

A Basic Tester Kit includes:

Density and Moisture Tester, Std. 8" Transmission, Dual Backscatter, NSN 6635-01-030-6896 (CPN Model MC-1)
Cast Aluminum Guideplate
Computer Tabulations
AC Charger
DC Charger (Batt. clips)
Sign Set
Phillips Screwdriver
Fiberglass Transportation and Storage Case (with key-alike padlock)
Tester Handle Key
Standard Drill Pin
Standard Block Assembly
Allen Wrench, Hex 5/32

Optional Accessories:

Clipboard Hammer

Unpacking

Damage upon receipt is the responsibility of the freight company. A claim should be filed with the freight company for damage. If the radioactive source may have been damaged, see Para 4-31.

1-1. GENERAL OPERATION

The tester is ready for use after unpacking.

There are no adjustments in the tester. All settings are factory set and sealed.

Batteries will be removed from tester and placed in the shipping case, whenever the tester is not in operation or is being shipped overseas. BATTERIES WILL NOT BE LEFT OR STORED IN THE TESTER. Refer to Section III for instructions.

If the batteries are low, a letter "L" will appear in the display. If the batteries are very low, the tester will "cutout". The "L" will appear, but the gauge will not operate. Recharge or replace the batteries (NICAD batteries only-NSN 6140-00-497-0490).

Charge the tester overnight when the batteries indicate "L" or low. The battery charger will not damage the batteries by overcharging.

A satisfactory program would be to charge the tester over a weekend once or twice a month depending upon use.

CAUTION

Due to excessive damage caused by battery leakage, only NICAD batteries, NSN 6140-00-4970490, will be used as battery replacement. Quantity of eight (8) each should be ordered.

WARNING

- * Do not attempt charging alkaline batteries. Alkaline batteries are not to be used in the tester.
- * Do not attempt charging zinc-carbon batteries. zinc-carbon batteries are not to be used in the tester.

Shipping Case

The shipping case is a fiberglass case with integral, molded foam padding (see Para 3-7 for shipping case maintenance).

Reference Standard The reference standard is a separate standard with the fiberglass case.

Reference standards are serialized for each tester, however, they are interchangeable between testers.

Guideplate

The guideplate is cast aluminum. It is entirely adequate for all surfacing and for support of the drill pin during transmission hole preparation.

Do not use the plate to remove the drill pin.

Drill Pin The standard drill pin is 3/4 inch diameter with a forged head.

1-2. MODES OF OPERATION

BACKSCATTER is the technique of placing source on the surface of a flat material. This technique measures only the top 2-inches of soil and is very sensitive to surface roughness or quality of site preparation.

TRANSMISSION is the technique of pre-drilling a small hole in the ground and then inserting the radioactive source in the ground via the movable source rod. This technique permits measurement to specific depths to 12-inches and is insensitive to surface roughness. It is very accurate and is the preferred method of measurement for soils.

DENSITY is measured by backscatter or transmission.

MOISTURE is measured by backscatter only. It is not a transmission phenomenon. The tester uses a separate source in the bottom casting to insure constant

backscatter operations and permit simultaneous counting of both density and moisture.

Approximately 90% of the moisture count is obtained from returned thermal neutrons from the closest 6 inches of soil.

1-3. PRINCIPLES OF OPERATION

General

Various elements, both naturally occurring (Radium) and reactor produced (Cesium and Americium) are unstable and are slowly decaying to a more stable state. The act of decay produces emissions of energy upon disintegration of the atoms.

These emissions are either "rays" of electromagnetic radiation (Gamma Rays) or are actual particles of material (neutrons, for example). Other emissions are produced from various radioactive materials; however, we are concerned with only the gamma and neutron radiation for purposes of nuclear soil testing.

Gamma radiation is emitted in several energy levels by a sealed Radium source and in a single energy level by a Cesium source. The Cesium level is 0.66 MEV and requires less shielding than the multi-level output of the Radium source. The fixed spectrum emission is also superior for soil density determination purposes. Cesium, a reactor produced isotope requires a license for use anywhere in the U.S. and in foreign countries.

Neutron emission occurs when an alpha particle emitter (Americium, Plutonium, or Radium) is mixed with Beryllium powder in a tightly compressed pellet. The alpha particles strike the Beryllium atoms to produce fast neutrons of an average energy of 5 million electron volts. The suffix "Be" is attached to the alpha source name to denote its use as a neutron source when it is mixed with Beryllium (RaBe, AmBe, PuBe).

These emissions are detected by appropriate detectors (Geiger Mueller tubes) for gamma and (Boron Trifluoride-BF3 or Helium-3 H3 tubes) neutron measurements. The resultant signals are displayed electronically as an index of soil density and moisture.

Radioactivity, both gamma and neutron, may be thought of as being similar to light from an incandescent bulb. The light rays diminish rapidly as we move away from the bulb (by the inverse square of the distance from the lamp), and they have the ability to penetrate various materials to some degree, ranging from nearly complete penetration (glass) to nearly complete blockage (metal shield).

Radiation obeys the same rules, although its penetration capabilities are generally much greater than light. The farther we are from the source, the safer we are, and the more absorbing material (shielding) we place between ourselves and the source, the safer we are. It is theoretically impossible to shield any radioactive source completely; however, careful tester design and appropriate choice of shielding materials can reduce the radiation to an acceptable level with negligible absorption by the user under proper operating procedures.

Gamma Radiation

Gamma radiation is electromagnetic "photon" energy capable of penetration of several inches of most materials. It is essentially high energy "light ray" energy. It is useful for the total mass measurement of heavy materials and is used to determine total density of soil.

Neutron Radiation

Neutron radiation consists of small, noncharged particles emitted from the source at an average energy level of approximately 5.0 MEV. This is known as "fast" neutron emission. Neutron detectors "see" only slow, or "'thermal" neutrons. Therefore, the fast neutrons must slow down or they will be ignored by the detectors. Neutrons slow down by colliding with other objects much like a rifle bullet ricocheting from rock to rock.

Collision of the fast neutrons with the nuclei of large atoms results in rebounding of the neutrons with little loss of energy. Collision with the orbiting electrons (approximately 1/1840th the weight of a neutron) produces little loss of energy. However, collision with an

object of the same mass will produce a major loss of energy or slowing down.

The only atom which can markedly slow down a fast neutron, and which we would likely see in soil, IS hydrogen. The hydrogen nucleus is the same mass as the neutron and slows down the neutron Immensely compared to collisions with other nuclei. The greatest loss of energy in any collision is when two similar mass objects collide.

A simple analogy is that of a golf ball colliding with a bowling ball. The golf ball would rebound with little loss of energy. The golf ball colliding with BB's (electrons floating around a nucleus) would push them aside. However, two golf balls colliding would produce a strong loss of energy in each of them, or a transfer of energy from one to the other.

This is what happens when a fast neutron hits a hydrogen atom. The neutron is markedly slowed down. A few collisions with hydrogen atoms reduces a fast

neutron to the slow or "thermal" energy at which the moisture detectors in the soil tester can "see" the slow neutron.

Thus, the moisture channel is in reality a "Hydrogen Analyzer" and is responsive to any form of hydrogen present whether it be in the form of water, or of some organic matter. It is possible to measure water on a construction site because the only form of hydrogen we normally see on a soil site is free water, the very feature we are trying to measure. However, bound water within the mineral matrix, organic matter, roots, or asphalt in an asphalt pavement would also provide hydrogen moderation and the neutron tester would "read" it accordingly. If we know the quantity of extraneous hydrogen, we can account for it in calibration and the tester can still be used for moisture determination.

Neutron radiation is emitted by any alpha producing source when mixed with Beryllium. CPN uses Americium/Beryllium (AmBe) source in its soil testers.

1-4. PANEL CONTROLS

Refer to Keyboard (Fig. 1-1) - Each Touchkey is outlined below.

	CLEAR	=	Stops count in progress. Stores this count and does not reset the display to zero. Cancels any other prior program keys (D or M-ONLY).
			Pushing a display touchkey will produce a display of the stored count. Note that the tester will go through its normal, rapid shutdown "beep" to alert the operator that the tester was stopped deliberately and not automatically.
NG)	LOCK	=	Interlocks TEST and STANDARD to prevent accidental field actuation of either o' these two functions. LOCK by itself will do nothing.
MODE SELECT - (PROGRAMMING)	STD	=	When pressed simultaneously with LOCK, will erase all four memories and will initiate a 4 Minute count into all four memories. This provides a check on the performance of the memories. The stored readings are normalized to the 1/4 minute time base for ease of operator computation. STD by itself will do nothing.
E SELECT -			The two STANDARD memories will be retained for future recall. Pushing any of the 4 OPERATION buttons will erase the standard count stored in the D-DATA and M-DATA memories and they will count normally.
MOD —	TEST	=	When pressed with any DISPLAY key, will light all 8's to test the display and the display command system. When pressed alone, it displays a random number which should be discarded.

OPERATIONS - (START) When pressed simultaneously with LOCK, will erase all four memories and will initiate a 1 minute count into all four memories. The displays will each read 65536 if the test signal has been properly processed. This tests all digital circuitry in the tester, but does not test the analog circuitry (high voltage supplies).

NOTE

After approximately Jul 78, TEST will count 1/4 minute and will display 16384. This shortens TEST time as a convenience.

D-ONLY = When pressed, the tester will count Density Only, retaining last moisture count. D-ONLY is cancelled by pressing an Operations Key or by pressing CLEAR.

M-ONLY = Same as D-ONLY except that only moisture is counted.

1/4 - MIN Initiates 1/4 minute count cycle. Use for Statistical check per Paragraph (STAT-CHECK) = 1-7.

1/2 - MIN = Initiates 1/2 minute count cycle.

1 - MIN = Initiates 1 minute count cycle.

2 - MIN = Initiates 2 minute count cycle.

PRECISION OF VARIOUS TIME PERIODS

(Read Para 1-7 through 1-9)

NOTE: ALL DISPLAYS NORMALIZED TO 1/4 MINUTE.

(in PCF @ 120#)	<u>1/4</u>	<u>1/2</u>	<u>1</u>	<u>2</u>
BS	1.20	.85	.60	.42
AC	1.00	.71	.50	.35
6"	.50	.35	.25	.18
Moisture	.50	.35	.25	.18

Legend:

PCF - Pounds Per Cubic Foot

BS - Backscatter

AC - Asphaltic Concrete

NOTE

Pushing any Operations key will immediately restart tester and will supersede prior command.

Upon initial operation after a period of nonuse, press the CLEAR key to clear out any unwanted commands.

Press any of the four OPERATIONS keys to start the tester for normal density and moisture counting. All displays will be normalized to a 1/4 minute count base for simplicity of computation.

At the completion of any count cycle, the tester will "beep-beep" rapidly for ten beeps.

During the count cycle, a pulsing "beep" will be heard

If the batteries are low, but still useable, an "L" will appear in the display window. If the batteries are below acceptable charge level, the "L" will display, but the tester will "Cutout" and will not operate. Recharge the tester overnight when the batteries indicate "L" or low.




Figure 1-1. Keyboard Configuration

1/4, 1/2, 1 or 2 = SIMULTANEOUS COUNT OF BOTH CHANNELS

TEST + ANY DISPLAY = DISPLAY TEST (88888)

TEST + LOCK = INTERNAL CHECK (65536) (16384 after 7/78)

STD + LOCK = STANDARD COUNT (4 MIN)

D-ONLY or M-ONLY = ONLY COUNT ONE CHANNEL

NOTE

Some testers do not have the D-only or M-only keys.

CLEAR = STOP COUNT (Does not reset to zero)

D-DATA or M-DATA = DISPLAY LAST FIELD COUNT

D-STD or M-STD = DISPLAY LAST STD or TEST COUNT

NOTE

All count periods are "normalized" to a common time base of 1/4 minute to reduce operator work in the field.

After approximately July 1978, Internal Check (TEST+LOCK) will count only 1/4 minute and will display 16384.

Both TEST/LOCK and STD/LOCK count into all four channels as a cross check for counting accuracy.

If nothing at all happens when the touchkeys are pressed, the batteries are completely discharged or are disconnected. See Para 1-1. (General Operation) and also Para 3-5. (Battery Replacement).

If a mistake is made in starting the count, clear the tester with the CLEAR key and restart the sequence. This will clear all stored commands. If an erroneous time cycle was pressed, it is necessary only to press the correct touchkey to restart on the correct time cycle.

To take a density count only and NOT take a moisture count, press D-ONLY. The previous moisture count will be stored. The D-ONLY command will erase at the end of the count cycle. M-ONLY follows a similar circumstance.

The tester will automatically take its own 4 minute standard count, if STD and LOCK are pressed simultaneously. All four memories are erased and the standard count is fed into all four memories See Para 1-7, Standard Count.

The tester may be internally checked by pressing TEST and LOCK simultaneously. The tester will count for one minute, providing a fixed time base pulse to all four memories. All four memories should read 65536 at the end of the count period. (1/4 minute and 16384 after 7/78).

The LOCK key prevents an accidental field erasure of the standard count memories by an inadvertent pressing of the STD or TEST keys.

WARNING

Insure that the tester handle is placed in the safe position prior to placing the tester into the approved carrying case.

Upon completion of testing, remove battery pack and store in storage case separately from tester. There is no ON/OFF switch and today's last readings will still be available when the tester is used next.

1-5. SOURCE ACTUATION

The shutter is released by pulling back on the handle latch pin and pressing down on the handle.

The shutter is spring loaded and closes automatically when the handle is raised to the SAFE position.

The operator should insure proper latch or handle position by visually checking the latch pin for full forward location at the desired depth.

Lock the source rod when in storage by pushing in the lock button on the end of the rod. Key required to unlock handle.

After repeated tests the shutter mechanism will become dirty with accumulated soil pulled into the chamber on the end of the source rod. The shutter will jam or become sticky. See Para 3-6 for shutter maintenance.

Do not force the source rod to open the shutter.

1-6.TO MEASURE OR TAKE A TEST, GENERAL

The following operations are required to take a field test:

- a. Take a Standard Count.
- b. Prepare the Site.
- c. Take a Reading on the Site.
- d. Determine Moisture and Density from the Charts.

e. Compute Percent Compaction.

WARNING

Insure that the tester handle is placed In the safe position prior to placing the tester into the approved carrying case f. Store the Tester Away.

1-7. STANDARD COUNT

Nuclear tester performance is affected by the decay of the radioactive source, by aging of the electronic components, and by minor mechanical wear and tear. As with any measurement tool, it is necessary to check the tester against a reference standard periodically to insure continued accurate operation.


This is achieved with your tester by the taking of a Standard Count on the reference standard furnished with the tester. Variations in tester performance will affect the Standard Count in the same proportion as they will affect the field counts.

Should variations in tester performance occur, these variations will affect both the field and standard counts proportionately, thus the final ratio remains constant and the tester remains accurate. The standard count thus becomes a means of continual calibration correction for the nuclear tester.

A further benefit can be achieved from the standard count due to the statistically reliable nature of the decay of the radioactive source over short periods of counting. The following paragraphs describe methods of evaluating tester performance using inspection of a series of successive counts in a standard count series.

1-8. HOW TO TAKE A STANDARD COUNT

Manual and automatic are two methods of taking a standard count. Both use the same mechanical setup. They differ only in the method of accumulating the final count results.



Place the source rod in SAFE position and be sure the tester is at least five feet from nearby objects.

Stand the shipping case on end and place the reference standard across the protective strips on the case Three buttons on the reference standard fit into accommodating depressions in the bottom of the tester. See Fig. 1-2.

At this point two counting methods may be used:

Manual, with successive counts.

Automatic, with four minute averaging.

The Manual method is used for periodic tester evaluation per Para 1-9.

The Automatic method is used for routine daily standard count accumulation into the tester Standard Count Memories.

Manual method

Successively accumulate 10 1/4 minute counts.

Do NOT use other time periods for purposes of evaluating tester performance under this program.

Proceed to Para 1-9.

Automatic method

Press STD and LOCK simultaneously on the keyboard and the tester will now clear all four memories, and will count a 4 minute count into all four memories. The 4 minute count will be normalized to a 1/4 minute common base for ease of subsequent computations.

Record the final average of either method on the daily field work sheet. The 4 minute automatic standard count will be retained in memory until erased by the next STD or by the taking of a TEST count.

1-9. TESTER EVALUATION USING MANUAL STANDARD COUNT

Periodically, the tester should be checked for good statistical operation. This is the ONLY way the operator can verify that his tester is operating correctly. Using this technique, it is possible to predict a future failure before it is severe enough to cause field errors or complete breakdown. This periodic evaluation should be performed at least once a month or whenever the tester has not been used for several months.

All nuclear method testers will vary in count rate from one undisturbed count to another due to the random nature of the radioactive source decay.

This is a very reliable random rate in that it faithfully obeys the laws of statistics relative to random sources of "population'...<u>where population = count rate</u> for nuclear testers.

Given: Any series of numbers taken among a varying population of events will exhibit a form of

standard deviation.

A series of nuclear tester counts, taken with the tester in an undisturbed position between all counts, will also exhibit a standard deviation around the average of the counts or mean.

Given: It is true that if the source of a series of events

is truly random, then the standard deviation of

the counts will equal the square root of the average count or mean.

Given

One standard deviation equals 68% of the population of events for a series.

Put these all together and:

+ 1 Std Dev =
$$-\frac{+}{\sqrt{\text{Average}}}$$
count = 68%Prob.

This finding can be put to good use in evaluating tester performance if we use it to inspect a series of counts taken in the standard count mode.

(The procedure may also be used to evaluate any series of nuclear tester counts. It is only necessary that the tester not be moved during the series of counts. Counts could be taken in any operating position, with another tester nearby, etc.).

Two evaluation methods will be described.

One uses a hand calculator with statistic functions, the other uses a manual checkoff method. Both techniques work well. Both should be used with ten successive counts taken in the 1/4 minute time period. (This will not work if counts are taken in other time periods due to the averaging or normalizing of the tester on long time periods).

1-10. CALCULATOR EVALUATION USING STANDARD DEVIATION

Enter the series of numbers into the calculator following calculator instructions for standard deviation. Determine the deviation and place it in one of the calculator's memories.

Pull up the mean of the series of counts and determine the square root of the mean.

Divide the deviation by the square root.

From the above discussion, if the series was "perfect", the square root would equal the deviation and the division would be "1.0". The series will seldom work out perfectly, however, and some variance from 1.0 will be observed. It is normal for this to lie between 0.75 and 1.25 with a general tendency towards 1.0 for a normal tester.

1-11. MANUAL EVALUATION

Take the same series of ten numbers. Use 1/4 minute time key.

If the standard deviation of the series is to lie within $\frac{+}{-}\sqrt{\text{average}}$ and this is to include 68% of the series, then this is merely stating that 68% of the numbers should lie within $\frac{+}{-}\sqrt{}$ of the average or mean and that 32% should be outside the average or mean.

That is, out of a series of ten numbers, 32% or about 3 out of 10 will lie outside plus or minus the square root of the average.

We simply add up a series of numbers, average them, determine the square root of the average and then add and subtract this square root to the average.

The resultant high and low limits will include 68% of the numbers in the series and 32% will be higher or lower than the limits.

The following example illustrates a typical tester placed in Standard Count configuration and using the Student Field Data Worksheet (DA Form 5448R (Moisture and Density Tester Field Data Worksheet)), Figure 1-3, to accumulate the data for ten counts. DA Form 5448-R is located at the back of this TM for local reproduction authority. This tester was normal in all respects.

We seldom are fortunate enough to have the series work out exactly 3 out of 10 each time we run a series. However, if we make an intelligent allowance for variations in numbers we will observe that the series will exhibit a general trend towards 3 out of 10. An occasional 2 out of 10 will be observed and an occasional 4 out of 10 will be observed. However, the statistic probability of a 5 out of 10 or a 1 out of 10 is very slim, and such splits should be very rare for a normal tester.

It is important to observe the average from one such Standard Count Evaluation to another. If the average is within approximately 1/3 to 112 of the square root of the prior average, then the difference between the two series is normal.

It is most desirable that the operator maintain a standard count diary with his raw data and his computations written down. This is most helpful to the repair shop should the tester require repair for some complaint of "drift" or "erratic" counts. Use DA Form 5449-R

(Moisture and Density Tester (Nuclear Method) Utilization Log) to record standard counts. DA Form 5449-R is located at the back of this TM for local reproduction authority.




Figure 1-3. Field Data Worksheet

Note that both standard count series produced a normal split of 3 out of 7 inside the limits of + square root of the average. A split of 2 out/8 in or 4 out/6 in is normal occasionally, provided the usual split is a 3/7 and provided the average standard count remains within approximately 1/3 square root of the prior series' average.

It is perfectly proper to use the Automatic Standard count for daily routine, but it is very desirable that the full evaluation be run periodically to ascertain that your tester is running properly, regardless of make or model.

1-12. WHAT DO POOR RESULTS MEAN?

If the tester consistently exhibits a tendency towards a high deviation, that is, the calculator method trends more towards the 1.25 end of the spread than the 0.75 end,

and the manual method trends more towards 4 and 5 out instead of 2 out, the tester is indicating a tendency towards erratic counts beyond the normal projected deviation to be expected.

This erratic tendency is likely to come from a dirty high voltage supply, dirty detector connectors, or a failing component in the high voltage supply. Simply dusting off the supply may eliminate the problems. If this does not work the tester may require repair.

If the tester consistently exhibits a tendency towards a low deviation, that is more towards the 0.75 end for the calculator or more towards a 2 out or 1 out for the manual method, then the tester is picking up some periodic count rather than purely random count, tester repair may be required.

This smooth tendency is probably due to a failing filter network in some supply system which would allow a steady noise to creep into the system, a noise not subject to the variation of a random source.

Using either of these techniques, either calculator or manual evaluation, the operator can maintain a close watch on his tester and can be confident that his field work is accurate without electronic error or decay drift.

It is not adequate to merely run the automatic standard count and not run the periodic evaluation.

The daily automatic standard count will serve to verify general good operation, but it will also mask minor variations between counts that are indicators of potential tester failure.

The standard count in the automatic mode can cover up erratic tendencies such as a noisy detector, or other problems. The tester may appear very normal under such conditions.

We must maintain a professional attitude about our equipment and its operation at all times.

This means frequent checking of standard count and tester performance.

1-13. SITE PREPARATION

Tests will only be as good as the quality of site preparation. The site must be flat, free of voids, with surface irregularities filled with native fines or sand.

If transmission is used, the hole must be neat and square to the surface, and protected from puffing or movement during the drilling of the transmission hole or retraction of the drill pin.

It is desirable to rotate the tester 900 or more on rocky soils to obtain the largest sample size practicable. This

avoids the possibility of a single rock skewing the results.

Site Selection

CPN advocates statistical testing wherein the average of several test sites is used to accept or disapprove the project. The nuclear tester is much faster than the older, conventional test methods and statistical programs may be employed.

Using the CPN Random Selection Cards, (See Appendix J) select five site locations and take one test on each location. Average the sites for a statistically valid evaluation of the project.

Each site must be free from major protuberances reasonably smooth, and should be representative of the overall project.

Preparation

Using the CPN Cast Aluminum Guideplate, smooth the surface, removing large stones or surface debris which would prevent the tester from sitting flat. Any technique of smoothing, pounding, or other flattening will be adequate. It is impossible for the operator to pound the site so hard as to compact the site. He can only flatten down the top 1/16" or so to provide a good surface.

Using a sieve, shake some native fines or sands over the site to fill in any remaining air voids to remove possible errors from surface roughness.

The tester must rest upon native promontories, however, with only the voids filled in. The tester must not rest upon a "cushion" of filler material.

If the test site is asphaltic concrete, (AC), a coffee can "salt shaker" with fine sand may be used to fill in the voids.

The tester has a reduced bottom surface to provide the best possible seating under adverse conditions. This facilitates seating the tester on AC where a rubber tired roller has been used.

1-14. SEATING QUALITY EXPERIMENT

An experiment can be run to test the required effort

to obtain good seating:

- a. Take a backscatter reading prior to any work.
- b. Perform a minimum of surface work, take another reading.
- c. Perform more work, clearing, smoothing, etc, and take another reading.

The readings will drop in count rate as the tester is seated better and better. When readings have ceased to change, the tester is seated as well as it can be. This is then an index of the work required to obtain good seating.

1-15. SURFACE ROUGHNESS ERROR

Surface roughness error is the error introduced into the tester readings due to the rough surface obtained during seating. Such error was exhibited in the work in paragraph 1-14.

The dual position tester backscatter feature provides a means of selecting a shallow backscatter position (AC Position) for best results on thin lift Asphaltic Concrete and a slightly deeper backscatter position (BS Position) for use on soils or thick Asphaltic Concrete.

Unfortunately, the thinner the tester attempts to read, the more sensitive it is to surface roughness.

The following test will illustrate the potential roughness error and will permit the operator to use the tester to his best advantage with a minimum of error.

- a. Take a reading in both normal backscatter (BS) and in the shallow backscatter (AC) on a smooth concrete surface. Use the 2 minute count period for this test.
- b. Raise the tester the thickness of a dime, approximately 0.050-inch, producing a 100% void under the tester.
 - c. Repeat the readings in both AC and BS test.
- d. The tester will read approximately 3 pounds per cubic feet (PCF) low in BS and will read approximately 7 (PCF) low in AC.

This will vary slightly between testers of the same design and between testers of different manufacturers.


The error is an index of the amount of effort required by the operator to obtain good seating with any specific gauge.

1-16. DEPTH OF READING (Fig. 1-4.)

Certain physics limitations present themselves in depth of reading of nuclear testers.

No matter how the design of the tester is made, soil testers using GM tubes and radioactive gamma sources of approximately 0.7 MEV energy will not read more than 5-inches deep for 99% of their response in construction materials from 100 to 170 PCF.

If the radioactive source is well collimated as in the BS position, the tester will tend to pick up 90% of its returned radiation from approximately 2.5-inches deep.






Figure 1-4. BS and AC Positions

NOTE

The tester's raised (collimated) BS position blocks shallow gamma rays to provide deeper reading for soils. Lower AC position permits utilization of shallow rays to decrease depth of effective reading for shallower materials. (Surface roughness effect increases with AC position, however, and site prep requires care).

If the radioactive source is NOT collimated, the source being positioned at the bottom of the tester permits gamma rays to stream at an angle backwards towards the detector, the tester will pick up 90% of its returned radiation from approximately 1.5 - 1.75 inches deep.

Both positions will read the same overall depth for 99% of its returned radiation.

Thus, if the intention of the reading is to read a soil embankment of more than 2-inches thickness, then either BS or transmission should be used to insure the deepest possible reading.

If the density of a thin lift overlay is desired or any lift of less than 2-inches thickness, then it would be desirable to use the AC position and to place the additional surface seating care required in these shallow positions.




Figure 1-5. Effective Depth-Inches

1-17. TRANSMISSION ROUGHNESS ERROR

Transmission is inherently an accurate mode of measurement in all respects. It is a linear measurement with depth and is not fooled by varying densities near the soil surface and is comparatively immune to surface roughness problems.

Seating area requirements for the transmission mode comprise only a 2×6 -inches wide surface under the GM tube, thus the tester can be used on very difficult surfaces where a backscatter tester would be impossible to use.

Transmission surface roughness error is approximately 1/10th that of the backscatter errors.

1-18. DRILL TRANSMISSION HOLE

The transmission hole must be vertical to the smoothed surface and must not be damaged in the process of drilling and retraction of the drill.

Place the Cast Guideplate on the surface and use the guide to support the drill pin. Hammer the drill pin into the ground, tap it lightly to loosen it, twist it to further loosen it, and pull it out of the ground.

WARNING

Do not use the cast guideplate as a retraction device. It will damage the top of the guide hole.




Figure 1-6. Transmission Hole

NOTE

During drilling and retraction, the operator's weight on the guideplate confines the soil to prevent eruption or puffing.

The operator's weight on the plate confines the soil in drilling and in retraction to produce a neat, undamaged hole for the readings.

On heavily compacted material above 95%, cracks may appear in the soil. These will not affect the readings if the tester is placed so no crack extends from the source directly back to the detector tubes.

If the drill pin "walks away" as a result of striking a rock in the soil, try another location. If repeated tries cannot produce a satisfactory hole, then the test should be aborted.

1-19. SEATING THE TESTER IN TRANSMISSION

From a radiation exposure standpoint, it is undesirable to have to spend a great deal of time "playing" with the tester in seating it. Locating the hole for insertion of the source rod in transmission is frequently a problem and CPN has made this easy for the operator.




Figure 1-7. Positioning Source in Prepared Hole

A small cast lip (Fig. 1-7) is on the front of the bottom casting to provide for insertion ease.

- a. Place tester on prepared site so that the source tube hole is positioned reasonably close to the hole in the surface to be tested.
- b. Standing slightly behind the tester with your feet placed comfortably apart, flex your knees and bend your body slightly forward above the waist.
- c. Brace your right elbow on your right knee, grasping the cast lip of the tester with the fingers of your right hand.
 - d. Put your left hand on the source tube handle.
- e. Allowing the left edge of the tester to remain in contact with the surface, lift the right edge of the tester two to four inches above the surface.
- f. Release the spring lock on the source tube handle with your left thumb and push the source tube down approximately 2 inches.
- g. Peeking over the front edge until you just see the end of the source tube, lower the right side of the tester and guide the source tube into the hole in the prepared surface.
 - h. Lower the source tube to the desired depth.

1-20. TAKE A NUCLEAR READING ON SITE

Once the site is prepared, a standard count is available, and the tester seated, a field count can be obtained.

No advantage is gained from repeating a count on exactly the same site. It is permissible to rotate the tester 900 or more on a site to obtain two views of the location, but there is nothing to be gained from taking more than one count without at least moving the tester. It would be better to move the tester to another site nearby for the second count in order to obtain a better representative sample of the overall project.

Reviewing paragraph 1-9 and applying these statistical factors to the tester readings, we will observe that:

a. The tester will not count exactly the same two times in a row except by coincidence.

- b. This count variation is a function of the random source decay.
- c. This will obviously produce a predictable error in the final result of PCF values.
- d. The error, identified as "Precision" can be charted as follows: (1 minute counts)

			±PCF
BS Regular	@	120 PCF	.60
BS (AC)	@	120 PCF	.50
6" Trans	@	120 PCF	.25
Moisture	@	15 PCF	.25

This error is the tester's repeatability. The final accuracy obtained on the site is also a function of how accurately the factory calibrated the tester and how carefully the operator chose and prepared his site.

The value of the final result is also a function of how representative of the overall project the selected site is.

Precision is also influenced by the selection of time period on the Operations Touchkeys:

(in PCF)	<u>1/4</u>	<u>1/2</u>	<u>1</u>	<u>2</u>
BS	1.20	.85	.60	42
AC	1.0	.71	.50	.35
6"	.50	.35	25	.18
Moisture	.50	.35	.25	.18

Select the time touchkey which will produce the desired precision versus time available for the test

1-21. FIELD COUNTS AND TRANSMISSION

- a. Site is prepared, standard count recorded
- b. Seat tester firmly.
- c. Release source handle trigger and lower source to BS or AC position as desired.
 - d. Press desired Operations Touchkey.

Tester will commence slow "beep" during count.

e. Tester will "beep-beep" fast at end of count.

Press DENSITY DATA or MOISTURE DATA key to read count from tester.

Record this on the Field Data Worksheet.

NOTE

DA Form 5450-R (Moisture and Density Field Test Sheets) (an abbreviated version of DA Form 5448-R) can be found at the back of this TM.

Moisture is taken at all times during any density count The moisture source is separately mounted within the tester. Only the density source is in the source rod. The moisture reading taken in the BS position is recorded on the worksheet.

f. If a density only or moisture only count is required, press the D-ONLY or M-ONLY keys to take only the desired count. The other prior count will be saved in memory.

1-22. FIELD COUNTS, TRANSMISSION AND MOISTURE

- a. Site is prepared, standard count recorded, hole drilled.
- b. Seat tester over hole and insert source per Para 1-19.
- c. Insure that handle is firmly in the desired depth slot.
- d. Firmly pull tester back against the hole to seat the source against the hole wall.
 - e. Depress the desired time key.
 - f. Record the data from density data displays.

1-23. DETERMINE DENSITY AND MOISTURE FROM CALIBRATION CHARTS

Para 1-7 describes the use of a ratio for determination of the final values, the ratio serving to cancel out long term drifts from source decay or component aging.

It is now necessary to record the field count data and to divide that data by the recorded standard count to obtain the desired ratios. The ratios will then be taken to the appropriate calibration charts to determine the final answers.

Calibration charts are provided by the manufacturer (CPN) with the tester. The charts are also titled to indicate the purpose of each chart; backscatter density, backscatter-moisture, transmission at "2", etc. The charts are also identified with a serial number which matches the serial number of the tester. Charts and testers are not interchangeable.

RATIOS	DENSITIES
0 246	127.00
0.243	127.50
0.240	128.00
0.238	128.50
0.235	129.00
0.232	129.50

Figure 1-8.

NOTE

The ratios listed in the left column cross over to the densities listed on the right. For example, the ratio 0.240 crosses over to 128, which is the density in pounds per cubic foot weight (PCF). If the ratio was calculated to be 0.239, which is not listed, the ratio would then have to be rounded up or down (usually down) to the closest listed ratio so the density can be determined.

The following portion of DA Form 5448-R (Fig. 1-9) provides data for a typical density test using a tester at three sites on a project and with three depths taken at each site. The average of the three readings will be used as the final evaluation density for the project.

The standard count is taken from Para 1-11.

FIELD COUNTS

STANDARD COUNT

= RATIO

1-24. DETERMINE DRY WEIGHT

Subtract the moisture in PCF from wet weight for each density depth to obtain a dry weight. The dry weight (DA Form 5448-R) is then used for the determination of Percent Compaction and Moisture (See Fig 1-10).

1-25. COMPUTE PERCENT COMPACTION

The tester has now provided the actual, in place density and moisture obtained during compaction of the soil Whether this is good enough for the project is not known until we compare this in place density with the soil maximum density produced from a laboratory sample of the soil being tested (found in a Corps of Engineers handbook or determined by actual laboratory tests). This is the "Proctor" value or soil maximum density. To obtain a Percent Compaction (DA Form 5448-R) we divide the dry density by the Proctor value which is assumed to equal 130 PCF. (See Fig. 1-11).

 $\frac{\text{DRY WEIGHT}}{\text{OPTIMUM}} = \frac{\% \text{ SOIL MAXIMUM DENSITY}}{\text{COMPACTION}}$

1-26. COMPUTE PERCENT MOISTURE

Percent moisture is expressed as a percent of the dry weight in soil testing. Therefore we must divide the moisture in PCF by the dry weight in PCF to obtain percent moisture dry, DA Form 5448-R (See Fig 1-12).

H20 IN PCF DRYWIEGHT = % MOISTURE DRY

1-27. COMPLETION OF TEST

Upon completion of the testing and before tester is placed in carrying case, remove battery pack and store it separately. It is not necessary to turn the tester off or perform other operations. Also insure that the tester handle is in safe position prior to placing the tester into storage case.

1-28. CALIBRATION CHARTS

CPN calibration curves or charts are determined from tests taken on CPN calibration standards.

CPN uses a variety of standards for gauge calibration including natural standards and artificial (metal) standards. Unfortunately, there is no "National Bureau of Standards" in the nuclear soil gauging industry. Each manufacturer does his best to provide an accurate set of standards and he checks these standards against other standards wherever possible. The major manufacturers generally agree quite closely on their standards although some disagreement may exist here and there with standards of users who have made them themselves.

If the chart produces results suspected of being in error (conflicts with field results of other testers or of other compaction measurements) do the following:

Check operation of tester. (See Para 1-7).

Check accuracy of other testers.

Retest a prior site that was correct

1-29. MOISTURE MEASUREMENT ERRORS

The moisture channel is actually a "hydrogen analyzer". It responds to moderated neutrons, the moderation largely a function of neutron collisions with hydrogen and only incidentally a function of collision with the large nuclei of other atoms.

Thus, any source of hydrogen will result in moderated neutrons and the tester will attempt to respond accordingly.

This can be used to our advantage in deliberately measuring asphalt content where the only hydrogen that we would expect to "see" would be from the hydrocarbons in the AC and not from moisture.

However, the hydrogen analyzer aspect of the tester can also be to our detriment if the soil we are trying to measure has hydrogen from other than free water around the mineral particles.

FIELD COUNTS	B/S DENSITY TRANS. 2"	TRANS. 4"	TRANS. 6"	TRANS. B"	B/S MOISTURE
	1470	8180			1236
	1463	8119			1237
	1448	8037			1242
TOTAL	4381	24336			3715
AVG -	1460	8112			1238
by StCt.	6076	6076			3101
ESULTANT RATIO	0.240	1.335			0.399
WET DEN PCF	128	127.5		PCF H	100 14.0

Figure 1-9. B/S Density and Moisture, 4" Transmission

DRY WEIGHT	(BS)	(2")	(4")	(6'')	(8'')
WET DEN PCF	128		127.5		
LESS H ₂ O	- <u>14</u>		- <u>14</u>		
DRY DEN PCF	<u> 114 </u>		113.5		

Figure 1-10. Dry Density PCF

PERCENT COMPACTION			
DRY DEN PCF	_114	113.5	
SOIL MAX DEN	130	130	
COMPACTION x 100 %	87.7	87.3	

Figure 1-11. Percent Compaction

PERCENT MOISTURE			
H ₂ O PCF	14	14	
DRY DEN PCF	114	113.5	
X 100 = % MOISTURE	12.3	12.3	

Figure 1-12. DA Form 5448-R (Percent Moisture)

Compaction is obtained by coating the surfaces of the soil particles with a thin, molecular film of water to provide lubrication during our compaction effort. This is the only moisture we wish to measure with the tester.

Unfortunately, however, some soils have other forms of water known as "bound water" or water of hydration, which bind themselves into the mineral matrix and become a part of the soil particle. This form of water does not contribute to the compactive effort, but is measured by the neutron channel in the soil tester.

A simple corrective technique can be used to cancel this error when detected.

The soil tester will always indicate a higher than actual moisture value when this error exists.

It will be advisable to make periodic checks with conventional testing means to verify that the problem continues to exist in a given embankment.

An infrequent form of error is occasionally observed in desert countries where a high deposit of boron may result in absorption of the moderated neutrons producing a low moisture reading. This is highly unusual. A compensating calibration curve must be prepared using oven samples for this occurrence.

1-30. MOISTURE REFLECTION ERRORS

Known as "trench wall effect", this error is the result of reflection of moderated neutrons from the surface of a nearby object to the tester, usually a trench wall or a backfill wall.

Knowledge of how the tester is constructed will minimize this error and a simple test can be conducted to determine the magnitude of this error in the soil tester.

The moisture detector is a long, 2 inch diameter tube lying along the long axis of the tester under the display area. Its most sensitive side is the side presented to the exterior of the tester on the display side.

The least sensitive area is the ends of the tube and the side of the tester facing the operator as he views the

display.

It is proper procedure to place the tester so the broadside of the tube "looks down" the trench rather than at the wall.

Conversely, the geiger tube used for counting is located at the end of the tester furthest from the source and lies across that end. It would be best that this end be placed looking away from a wall for density tests.

Thus, good practice in a confined space would be to face the tester longitudinally in the direction of the trench for density testing and across the trench for moisture determination.

The operator can make a simple experiment with his nuclear tester to determine the effect of the wall reflection.

Set the tester in the center of a concrete floor, at least ten feet from nearby objects. Determine a moisture reading on that spot.

Fill a two cubic foot box with damp sand or soil and slowly move this box towards the tester on each of the four sides, about 6-inches at a time, taking readings each time. Note where the box is when the readings produce an error sufficient to cause an unacceptable field density conclusion.

This distance from the tester will differ on the four sides, and will differ between density and moisture.

In general, the tester can be used in a trench of not less than 18-inches span, preferably about 24-inches wide, for best results on moisture. The density readings can be taken with the tester nosed directly into the wall, but the geiger tube end must be kept at least six inches from the wall.

Conduct this experiment with your testers to determine your own limits in your shop.

1-31. ASPHALT CONTENT DETERMINATION

The tester's moisture channel is a hydrogen analyzer.

Use this to good advantage by extending the use of the tester to become a good AC content tester for field use.

Some additional hardware is required

- a. Metal pan, approximately 16" x 16" x 4".
- b. Compaction plate...5" x 5" x 1/2" with a rod welded to it and a sliding weight to slide on the rod and permit a controlled compaction effort on the plate.
 - c. Accurate scale to approximately 30 pounds.
- d. Straight edge and a finely graduated metal ruler. .. C-clamp to clamp the ruler to the straight edge. The straight edge should be approximately twice as long as the pan's longest dimension.

Procedure

Establish an initial calibration curve in the laboratory. After that, the same procedure is used in the field to establish the desired asphalt content value.

Prepare a sample of asphaltic concrete at a specific weight each time. This weight will be selected at the outset and will be used for the initial calibration as well as for all subsequent field tests.

This weight should be approximately 95% of optimum for the typical asphalt weights used in the user's locale. Round it off to a practical value for ease of preparation and calculations.

Initially prepare a sample at a known density, evenly compacted in the pan in one inch thick layers. Thereafter prepare the sample in the same one inch layers for all field work using the same technique exactly.

a. Measure the pan and determine precisely the volume in each inch for the bottom three inches. The fourth inch will merely become a catch basin for the light material that is compacted into the third inch.

If a 16" square pan was used, then each inch would hold precisely 0.1481 cubic feet.

Establish a density to be compacted in the pan each time. . this is 135 pounds/cubic feet, a value that is about 95% of optimum for the region that is going to use this test. It will vary, of course, for different parts of the world.

If the desired density is 135 PCF, and if each inch in the pan represents 0.1481 cubic feet then a one inch layer compacted to 135 PCF would actually weigh 19.99 pounds.

135 PCF x 0.1481 CF = 19.993 pounds.

b. Measure out exactly 19.99 pounds of hot AC mix and dump this into the pan.

Spread it around evenly and then compact it down to a depth of exactly one inch.

A 1/16th inch error in height is equal to a 6% error in final compaction. . an error of 8.1 PCF, something not accepted or calibration work.

c. Compact a second and third lift in the same manner.

Weigh the final pan of material and it should weigh exactly $3 \times 19.99 = 59.97$ pounds plus tare.

- d. Set the pan atop the tester shipping case in the same manner as you would take a standard count. Remove the reference standard from the case.
- e. Take several tester moisture readings on the compacted surface, revolving the tester 900 each time. Record the readings.
- f. Run an extraction of a sample of the compacted material to determine asphalt content.
- g. Plot the tester reading (in ratio) on the vertical axis of a calibration sheet similar to that used fur the CPN moisture curve, and plot the asphalt content PERCENTAGE across the bottom.

This curve is now an accurate asphalt content curve.

Zero is obtained by compacting a bone dry

sample of the material into the pan and taking a tester reading. It may not be possible to obtain 135 PCF dry material In the pan, however. Any error resultant from an inadequate compaction of the dry material will be negligible.

h. This curve IS accurate as long as the same 135 PCF is produced in the field. If other weights are desired for field compaction, then a separate curve should be produced for each desired field weight. Obviously, It is desirable to attempt to select only one or two appropriate field weights to minimize curves and effort

The pan can also be set on legs, anything at least two feet from the ground and five feet from surrounding objects Watch out for the tailgate of a pickup truck due to the presence of the gasoline tank and spare tire, both heavy in hydrogen.

1-32. ROOF MOISTURE TESTING

The soil tester becomes a most useful tool when it is used to measure the trapped moisture in a built-up roof. It is always difficult to locate a leak and to estimate the area of roof damaged by a leak Built-up roofs generally consist of a waterproof membrane sandwiching a thick layer of insulation between it and the roof structure If moisture becomes trapped In the insulation, the insulating value is damaged and the roof structure is subject to deterioration, dry rot, or other injury.

Any attempt to puncture the roof to investigate the degree of moisture intrusion is undesirable because of the additional damage the testing introduces. The non-destructive nature of the tester provides for rapid, accurate determination of roof moisture without penetration of the roof.

1-33. ROOF MOISTURE IS A RELATIVE MEASUREMENT

All roofs are not constructed alike. They vary in thickness, type of insulation, type of underlying construction, and type of waterproof membrane.

Some will have hydrocarbons associated with the construction materials and Insulation, others will have little Internal hydrocarbon structure. The moisture channel is affected by the internal hydrogen in the same

manner as it is with bound water in soil. Thus, a certain background reading will be present even though the roof is "dry".

A roof with wet insulation 4 inches thick will provide a higher moisture count rate than will a roof with the same wet insulation, but only 2-inches thick.

Thus, the reading of a moisture tester on a single location on a roof cannot be used alone as an index of whether the roof is wet or dry.

A series of readings over a representative sample of the roof must be taken and readings compared to each other to produce a profile which can then be interpreted as a moisture profile.

1-34. PROCEDURE

Starting from a corner, take readings on a 10' grid pattern over the roof. 1/2 minute readings should be quite adequate. Record the actual counts Do not bother with computing a ratio and do not attempt to use the soil moisture chart to determine actual moisture in engineering units.

It will be noted that a majority of the readings will fall at some minimum value for this particular roof This will correspond to the dry readings. Other readings will be higher. The highest readings should be further explored with additional readings taken on 5' grid to establish the boundaries of the higher readings.

These highest readings represent the wettest portion of the roof It may be necessary to make a test penetration to determine the actual moisture present at this highest value By observation this can be classified as an arbitrary "moist", "wet", or "saturated" value.

Readings in between the dry bottom readings and the moist high readings can be classified proportionately.

Extra readings should be taken around chimneys, vents, stairwells, cornices, and other protuberances. Wet areas should be tracked to their perimeters.

1-35. PLOT ROOF MOISTURE PROFILE

On a drawing of the roof, plot the observed values

as dry, moist, wet, and saturated for each test location.

Draw "contour" lines between the values to enclose each given value area. A moisture pattern is immediately apparent.

For additional clarity, color the enclosed areas with a specific color pencil to designate the various moisture values.

Intelligent action can now be taken as to where repairs should be commenced on the roof and over what area. It may be obvious where the moisture is coming from... flashing, a previously unobserved puncture, crack, or low spot, once the pattern is drawn.

1-36. RADIATION SAFETY

Density is not a factor on the roof, therefore the shutter will be kept closed and gamma radiation beneath the tester is negligible.

A typical industrial built-up roof will be more than 8' above head height of occupants of space beneath the roof. Neutron radiation at that distance is less than 0.025 mrem/hr. Since the tester is constantly being moved from test spot to test spot, there is no radiation danger presented to occupants of the building during testing and no special provisions should be taken for safety.

The operator will take many readings in a day, however, and good practice dictates stepping back from the tester 5' or more while the tester is counting.

1-23/(1-24 blank)

SECTION II

RADIATION INFORMATION

WARNING

Always wear your beta-gamma badge or TLD badge when operating the tester and upon entering tester storage areas, and at other times as directed by the Local Radiation Protection Officer (LRPO).

NOTE

There is no danger to the Army Community and the general public when the tester is used, transported, stored, serviced and leak tested according to Section IV of this manual. Normal radiation exposure from a properly operating tester is far less than the maximum allowed dosage of 1.25 rem per calendar quarter, on an accumulated calendar year dosage of 5 rem.

2-1. GOOD RADIATION PRACTICES

- a. Always keep in mind how close you are to the tester when it is operating. Distance from the tester is your best precaution.
- (1) At any given distance, no matter how far or how close to the tester, when that distance is doubled going away from the tester, radiation exposure gets four times weaker.
- (2) At any given distance, no matter how far or how close to the tester, when that distance is cut in half going toward the tester, radiation gets four times stronger.
- (3) Always stand at least 5 feet away from the tester when it is operating.
- (4) Complete taking measurements as quickly as possible.
- (5) You are authorized to carry the tester only by its handle or in its carrying case. In either instance, the handle must be locked in the SAFE position.

- b. A busy workday can result in as many as 30 measurements. A busy workweek could result in 5 days of extensive measurements.
- (1) Even with extensive use, the dose to be expected from a properly operating and used tester is only 1/400 of the maximum allowed dose.
- (2) Thirty (30) measurements per day times 10 seconds per measurement equals 300 seconds or 5 minutes per day of exposure at 2 feet from the tester.
- (3) Five (5) days times 5 minutes equals 25 minutes, rounded off to 30 minutes.
- (4) One-half hour times 1/2 mrem/hr equals only 1/4 mrem exposure during a busy workweek.
- (5) Operators are allowed an average weekly dosage of 100 mrem.
- (6) Keep the curious away when you are on the job, but don't frighten people by making it a big thing.
- (7) Always place the handle in the SAFE position when the tester is not in use.
- (8) Store the tester under lock and key when not in use. Only radiation protection officers and operators shall have access to tester and carrying case keys.
- (9) Make it a habit to routinely wear your betagamma film badge or TLD badge. Store the badge at an approved location.
- (10) Do not intentionally expose the Cs 137 source to air by extending it out of the tester in the 2, 4, 6 or 8-inch transmission positions. Do not withdraw the source from the measurement site before returning the handle from the SAFE position.
- (11) Work fast and keep proper distance from the tester.

2-2. COMMON RADIATION TERMS AND VALUES

Radiation is similar to light. It increases in intensity by a factor of FOUR each time the distance from the source is cut in HALF.

The radiation level drops by a factor of FOUR each time you move TWICE as far away from the source.

Certain terms are used to describe radiation factors important to tester users. Be familiar with the following terms and values:

CURIE: A term used to describe the size of a radioactive source. It represents a quantity of material disintegrating at the rate of 3.7 x 10 disintegrations per second, or the same rate as one gram of Radium. This is not an index of how dangerous the source might be, but only an index of quantity of the material in question.

MILLICURIE: One thousandth (1/1000) of a curie.

MICROCURIE: One millionth (1/1, 000, 000) of a curie.

ROENTGEN: A term describing the amount of radiation accumulated or exposed to, by standing near a large radioactive, unshielded source for a short time or near a small, unshielded source for a long time.

REM (rem): This is a better term for measuring human exposure accumulation than Roentgen because it has been corrected to provide a common base for radiation effects on people. Some radiation is highly penetrating and would be more potentially dangerous than other forms. The descriptions become equal when they are all corrected to the common rem base.

MILLIREM (mrem): One thousandth (1/1000) of a rem. Tester radiation levels are commonly measured in these very small units.

MILLIREM/HOUR (mrem/hr): A term used to describe the "brightness" of a radioactive gamma source. It is the strength of the radiation field at the point of measurement. This term is similar to footcandles of light when discussing light.

The brightness of the radiation field will be dictated by the type of radioactive material involved, the size of the source, the amount of shielding present, and the distance from the source. The total amount of radiation accumulated would then become a factor of how much time was spent in the radiation field.

NOTE

Since the tester uses small, well shielded sources, operators will be involved with only millirems of radiation and with levels which are only in the mrem/hr range.

100 mrem: Weekly allowed dose (5 rem/yr is max annual work dose).

5 mrem/hr: Average radiation dose at tester surface.

1/2 mrem/hr: Average gamma dose at 2' from tester (arm's length).

1/3 mrem/hr: Average neutron dose at 3' from tester (source to midtrunk distance when carrying tester).

1/2 mrem: Average heavy workweek accumulation of radiation for a tester user (1/200th of allowed dose).

SECTION III

OPERATOR MAINTENANCE PROCEDURES

WARNING

Do not attempt tester repairs. A faulty tester must be returned to a designated manufacturer's repair facility.

NOTE

- This section applies to basic maintenance to be performed by the field operator with a minimum of tools and equipment and with a minimum electronics knowledge.
- Maintenance covered by Section III involves only the use of common hand tools.
- Report any defects with the tester to the LRPO.

3-1. GENERAL

The tester requires little maintenance other than an occasional cleaning of the shutter and replacement of batteries.

- a. Service problems will occur In three areas:
- (1) Batteries.
- (2) Mechanical jamming of shutter needs cleaning.
 - (3) Electronic parts failure.
- b. CPN testers minimize service problems in these three areas by careful design: (1) Batteries are replaceable with NICAD batteries, NSN 6140-00-497-0490.
- (2) The shutter mechanism uses a quick release cleanout plate for ease of cleaning.
- (3) Electronics use CMOS circuits for long life and low current drain.

3-2. BATTERIES

The tester uses NICAD rechargeable 4 Ampere Hour D Size cells as standard batteries.

CAUTION

Due to excessive damage caused by battery leakage, only NICAD batteries, NSN 6140-00-497-0490, will be used as replacement. Quantity of eight (8) each should be ordered.

3-3. CHARGING (Fig. 3-1.)

Charge NICAD batteries only. Do not use alkaline or zinc-carbon cells!




Figure 3-1. Charging

NOTE

Optional 12 VDC Cord illustrated.

Plug the wall charger into 110/60HZ supply and plug the charger plug into the charger terminal on the tester. The tester batteries will automatically charge with no danger of overcharging.

CPN recommends charging over a weekend or overnight whenever the "L" low battery indicator is on.

NICADS lose approximately 1% of their charge per day through self-discharge. The tester will require charging if it has been in storage for a period of months.

When the tester is not in use, remove the batteries. Batteries are not to be stored in or left in the tester.

3-4. ANTICIPATED TESTS PER CHARGE

Current consumption is low due to use of CMOS circuitry. The total time between charges will be a function of the number of tests taken and the total time passage between charges.

A tester uses approximately 80 milliamperes during count periods. It uses approximately 2 ma during quiescent periods (no counting).

The NICADS have a 4,000 milliampere charge life when new and fully charged. (4.0 Ampere hours).

Thus, if the tester was not used during a charge period, the tester should operate approximately 2,000 hours per charge or approximately 83 days.

If the tester is used continuously with no rest, it would last approximately 50 hours and would provide 3000 one minute tests (50 x 60).

How long it will last for a given user and how many tests will be available, is thus a combination of passage of time and tests performed.

The charge should last for many weeks for a typical user.

3-5. BATTERY REPLACEMENT (Figure 3-2.)

Remove battery cover.

Remove the knurled nut retaining the battery pack in the tester.




Figure 3-2. Battery Replacement

Slide the battery pack out of the tester.

Remove the appropriate hold down plate atop the cell that is to be replaced.

Replace the cell with positive end (button end) UP.

Replace the hold down plate, replace the pack, and replace the battery cover.

Make sure the cover makes contact with the battery.

3-6. SHUTTER MAINTENANCE (Fig. 3-3.)

The shutter is a carbide-tungsten-lead combination for maximum shielding and maximum longevity




Figure 3-3. Shutter Maintenance

NOTE

Two 6-32 fasteners simplify plate removal.

It will require occasional cleaning due to dirt buildup within the shutter chamber.

Set the tester on its side or tail with the source in safe position. Stand behind the tester, peek over bottom of tester to see cleanout plate. Remove the 1/4 turn fasteners holding the cleanout plate in position.

Remove the plate and the captive shutter. The shutter and cleanout plate are an integral assembly. Do not touch the brass ring around the hole in the plate. Clean the shutter and the plate thoroughly, and clean out the shutter chamber with a brush (standing behind the tester while this is done). It is not necessary to extend the source to clean the end of the rod! Leave it alone!

Spray the shutter, cleanout plate, and the interior of the chamber with molybdenum-disulphide lubricant (Appendix E, Item 2). This lubricant should then be allowed to dry for a period of four hours or more to cure to the metal. This will provide a long life lubrication of the shutter for

maximum utility between cleanings.

NOTE

Do not force the source rod handle when the shutter is dirty.

CPN does not recommend the use of oil or grease for lubrication, however, several customers have reported good performance by spraying WD-40 (Appendix E, Item 3) into the shutter chamber without disassembling it. The use of WD-40 will not injure the tester or affect test results. If it works for you, use it.

3-7. SHIPPING CASE MAINTENANCE

The shipping case is a fiberglass case built to MIL Spec standards. It is very strong and will produce excellent service.

No special care is required on the case other than an occasional cleaning (both inside and out) with soap and water.

3-8. TESTER EXTERIOR MAINTENANCE

Wash the tester with soap and water if it becomes unusually dirty.

WARNING

Drycleaning solent PD-680 is toxic and flammable. Wear protective goggles and gloves and use only in a well ventilated area. Avoid contact with skin, eyes, and clothes, and don't breathe vapors Do not use near open flame or excessive heat. The flash point for Type #1 Drycleaning Solvent is 1000F (380C) and for Type #2 is 1380F (500C) If you become dizzy while using cleaning solvent, get fresh air immediately and get medical aid If contact with eyes is made, flush your eyes with water and get medical aid immediately.

Remove asphalt with PD-680.

Do not scratch display window with abrasive cleaners.

If a major accident damages the tester by puncturing, refer to Para 4-31.

3-9. DIAGNOSTIC QUESTIONS

- a. Does the display work? Are the display screws (4) tightened properly? On all channels?
 - b. Does the tester beep during counting?
 - c. Does the tester count in TEST MODE? What does it indicate upon completion?
 - d. Does it count in STANDARD MODE?

Is this a normal standard count?

What is the spread of the standard count? Refer to Section 1, Para 1-7.

e. Does it count in the field count conditions?

In all time sequences?

- f. How long has the problem existed? If intermittent, describe when it occurs as opposed to when it does not occur.
- g. The operator and LRPO can use diagnostic questions.

3-10. REPLACEMENT OF FUSE

The fuse is on the battery pack assembly.

A spare is also supplied on the assembly.

Replace fuses with the same size as removed, 1.5 Amp Slo-blo.

Repeated fuse blowing indicates internal electronic trouble.

SECTION IV

SAFETY PROCEDURES

GENERAL

- **4-1. PURPOSE.** This section provides life cycle controls for the Campbell Pacific Nuclear Model MC-1 Nuclear Method Tester, NSN 6635-01-0306896. Areas of responsibility are then identified and assigned for personnel involved in the life cycle program.
- **4-2. SCOPE.** These safety procedures apply to Active Army, US Army Reserve, Army National Guard, and all other installations, activities and units covered under the conditions of the Nuclear Regulatory Commission (NRC) license issued to the US Army Tank-Automotive Command (TACOM). Areas of involvement include procurement, storage, possession, shipment, transfer, operation, use and disposition of the tester.
- 4-3. LICENSE. NRC license number 21-01222-05 has been issued to TACOM. This license authorizes TACOM to purchase, use, and store the tester. Conditions of the license permit the tester to be issued throughout the Army without specific licenses being issued to individual users. The license also states how radioactive material shall be used to measure the moisture content and density of soils and asphalt. A copy of the license is furnished with each tester. Additional copies of the license may be obtained from the Commander, TACOM, ATTN: AMSTA-CZ.
- **4-4. REGULATIONS**. Refer to Appendix A for applicable Army regulations (AR), technical bulletins (TB) and codes of Federal regulations.

RESPONSIBILITY

- **4-5. PROGRAM MANAGER**. The Commander, TACOM, is the program manager. Areas of responsibility are:
- a. Ensure compliance with Section IV of this manual by monitoring reports of radiation leakage tests, incidents and equipment improvement. Utilize feedback reports from logistic assistance teams, inspection teams and reports from major

- Army command Radiation Material Control Points (RMCP).
- b. Maintain overall National Inventory Control Point (NICP) functions by managing the tester as a major item
- c. Provide coordination for tester disposal in accordance with AR 385-11.
- **4-6. NICP ITEM MANAGER.** Areas of responsibility are:
 - a. Account for each tester by serial number.
- b. Issue testers to specific units, based upon plans developed individually, or together, by the Chief of Engineers and the Commander, TRADOC.
- c. Ensure that the consignee for the initial issue of the tester is the commander of the specific unit (see b above). Under no circumstances shall the consignee be the Property Book Officer (PBO), Director of Industrial Operations (DIO), or the Transportation Officer (TO).
- d. Ensure that all requests are channeled through the using units RMCP, and that the requests contain certificates and appointing orders on the using Command Local Radiation Protection Officer (LRPO), Alternate LRPO and operator. The requests must also contain a proposed storage location drawing with background readings as per Section IV of TM 5-6635-386-12.
- e. Issue disposition instructions and coordinate disposal action in accordance with AR 385-11.
- f. Request documentation from manufacturer of testers received for overhaul or repair from user unit.
- g. Request documentation from manufacturer of testers in transit to user units.

- h. Provide the TACOM Safety Office (AMSTACZ) a listing semiannually (April and November) of where all the testers, by serial number, are located.
- i. Coordinate with the TACOM Safety Office (AMSTA-CZ), before release of testers from the manufacturer repair facility or Sharpe Army Depot or before transferring testers among using units.

4-7. COMMANDER OF MAJOR COMMAND. Areas of responsibility are:

- a. Ensure compliance with applicable regulations and Section IV of this manual.
- b. Establish at least one RMCP at the command level in accordance with AR 385-11.
- c. Appoint a Radiological Control Officer (RCO) for each RMCP. Ensure that the RCO is well informed concerning applicable regulations and Section IV of this manual (see Appendix A).
- d. Ensure that each installation and activity that uses the tester has an effective radiation protection program.
- e. Ensure the implementation of Title 10, Code of Federal Regulations (CFR), Parts 19, 20, and 21.

4-8. RADIATION CONTROL OFFICER (RCO). Areas of responsibility are:

- a. Review and approve the qualifications of the Local Radiation Protection Officer (LRPO) and the Alternate Local Radiation Protection Officer (ALRPO). Forward LRPOs, and ALRPOs or tester operator (for tester assigned units without LRPOs and ALRPOs), certificates and appointing orders, for approval, to the Commander, TACOM, ATTN: AMSTA-CZ. Also semiannually (April and November) provide the TACOM Safety Office (AMSTACZ) a listing of LRPOs, ALRPOs or operators (for tester assigned units without LRPOs and ALRPOs) for each tester using unit.
- b. When a qualified LRPO orALRPO, one or the other (assigned to the using unit or covering for the using unit), cannot be on-site within a four hour response time, the RCO shall suspend use of the tester until a qualified operator and LRPO or ALRPO are available. Any changes to LRPOs and ALRPOs or tester operator (for using units without LRPOs and ALRPOs) shall also be provided to fester until a

- qualified operator and LRPO or ALRPO are available. Any changes to LRPOs and ALRPOs or fester operator (for using units without LRPOs and ALRPOs) shall also be provided to AMSTA-CZ, TACOM.
- c. Maintain records in accordance with AR 385-11.
- d. Ensure that the tester is properly handled, as specified in the operator's manual, applicable regulations and Section IV of this manual. Periodically inspect and audit the records of installation and activities that have the tester.
- e. Ensure that radiation incident reports are submitted by electrical means to the Commander, TACOM, ATTN: AMSTA-CZ, not later than 24 hours after the incident has occurred.

4-9. USING UNIT COMMANDER (UC). Areas of responsibility are:

- a. Supervise implementation of the local radiation safety program in accordance with Federal and Army regulations.
- b. Ensure compliance with appropriate instructions, regulations and Section IV of this manual.
- c. In writing, designate the LRPO and ALRPO. Forward a copy of the appointing orders and the certificates of training. The LRPO/ALRPO shall be commissioned officers, non-commissioned officers, or civilians who have completed one or more of the following courses.
- (1) Radiological Safety Course 7K-F3, US Army Chemical School.
- (2) A training program approved by the RCO and the TACOM Safety Office, ATTN: AMSTA-CZ. A minimum of 40 hours.
- (3) USAREUR Local Radiation Protection Officers Course. A minimum of 40 hours.
- (4) Army National Guard Radiation Protection Officers Course. A minimum of 40 hours.
- (5) Operational Radiation Safety Course 4J-F2, US Army Chemical School.
- (6) Combination of the Calibrator Custodian Course and 51G Course (Material Quality

Specialist Course).

- (7) A minimum of 40 hours formal training consisting of:
- (a) Principles and practices of radiation protection.
 - (b) Biological effects of radiation.
- (c) Radiac instrumentation and monitoring techniques.
- (d) Mathematics and calculations that are basic to the use and measurement of radioactivity.
- (e) Operation and use of the tester (CPN Model MC-1).
- d. Establish a standard operating procedure (SOP) for use by using unit personnel that has been coordinated with the LRPO and RCO.
- e. Suspend use of tester if no qualified LRPO or ALRPO is available. Assign the tester operator to maintain and wipe test the tester in accordance with the TM, until trained LRPOs are available. LRPO to pertorm leak tests, radiation surveys to ensure the tester is used only by trained personnel, and to prepare and submit necessary reports in accordance with Army and Federal regulations.
- f. Notify the Commander, TACOM, of any radiation safety defects or hazards associated with the tester, and other actions that do not comply with Title 10, Code of Federal Regulations, Part 21.
- g. Requisition the tester from the NICP item manager in accordance with AR 385-11.
- h. Obtain prior approval from both the RCO and the NICP when transferring the tester from one unit to another, even within the same command, and when requesting disposal instructions. Units in OCONUS shall also obtain prior approval from their Radiological Control Office and the NICP, when transferring the tester between units.
- i. Notify the RCO, in writing, of any changes in LRPO and ALRPO assignments or tester operator assignments (if the using unit is without an LRPO and ALRPO).

- **4-10. LOCAL RADIATION OFFICER (LRPO).** Areas of responsibility are:
- a. Advise the unit commander on all radiation matters and ensure that all testers under the commander's jurisdiction are properly used and stored.
- b. Instruct personnel in subjects relating to safe work practices, emergency procedures, the harmful effects of radiation overexposure, and other required topics in accordance with Title 10 CFR Parts 19 and 29 CFR 1910.
- c. Write and update the SOP for the units radiation protection program. The SOP shall include procedures covering the tester.
- d. Report any radiation incident, accident or theft immediately to Commander, TACOM, ATTN: AMSTACZ, by the most expeditious methods available; i.e., telephone and teletype. Then, make the same report through channels, to the Surgeon General. Thefts will also be reported by the Commander, TACOM, ATTN: AMSTA-CZ, to the NRC. NRC Regional office telephone numbers are provided in NRC Form 3 which is part of AR 385-1 1. Information copies of reports shall be provided to all intermediate headquarters.
- e. Perform radiation leak tests in accordance with the procedure provided in paragraph 4-27. Leak test results shall include the LRPOs name and telephone number, and the tester serial number.
- f. Upon receipt of the tester, conduct the initial radiation survey of the storage area and surroundings. Make certain that a copy of the NRC license and other documents required by 4-19.a are appropriately posted at the outside entrance to the storage enclosure. A copy of the initial radiation survey shall be sent to the Commander, TACOM, ATTN: AMSTA-CZ. Periodic surveys conducted on at least a quarterly basis shall be recorded locally. Whenever a change in storage conditions occur, a copy of the survey shall be sent to the Commander, TACOM, ATTN: AMSTA-CZ.
- g. Ensure correct shipment of radioactive items, both to and from the unit. See Para 5-16.

- h. Review dosimetry records maintained by the medical facility in accordance with AR 40-14.
- i. As directed by Title 10 CFR Part 20 and in accordance with AR 40-14, limit radiation exposure to as low a level as can be reasonably achieved. Ensure that an ALRPO has been appointed, and that unit personnel comply with radiation standards.
- j. Notify the UC and Commander, TACOM, ATTN: AMSTA-CZ of any radiation safety defects or hazards associated with the tester, and with any other actions or circumstances that do not comply with Title 10 CFR Part 21.
- k. Ensure that records are maintained for each tester in accordance with AR 385-11.
- I. Ensure that all radiation survey instruments used for monitoring the tester are calibrated in accordance with TB 43-180.
- m. Be knowledgeable of the subject matter in this manual.
- **4-11. ALTERNATE LOCAL RADIATION PROTECTION OFFICER (ALRPO).** The ALRPO shall assist the LRPO as required.
- **4-12. OPERATOR**. Areas of responsibility are:
- a. Upon initial assignment, all operators shall be checked out by the LRPO to ensure that they can safely and correctly handle and operate the tester. Tester operators shall have attended and successfully completed a course on the care, maintenance, and operation of the tester. Training courses shall Include one of the following:
- (1) Material Quality Specialist Course, (51G), Ft. Leonardwood, MO School. By agreement with the school, it is possible to attend the three-day portion of the course that is related to operating the tester.
- (2) Tester Training Course. The course is taught by an instructor who has been certified by the New Equipment Training School, TACOM Maintenance Directorate or by the Ft. Leonardwood, MO training school. The instructor must also be approved of, by the TACOM Safety

- Office (AMSTA-CZ).
- b. Maintain the tester user's log. A copy of DA Form 5449 can be found at the back of this manual.
- c. The tester is under the control of the operator at the job site. The operator shall not leave the job site until the tester is secured and locked in its storage case. The operator shall then ensure that the tester is returned to its proper storage area.
- d. Secure the tester when it is not in use. Ensure that the tester is not used when there is no LRPO/ALRPO assigned to the unit.
- e. Be knowledgeable of the subject matter in this manual. Supplemental operating, maintenance, and repair parts instructions are included in Section V of this manual.
- f. Notify the LRPO of any radiation safety defects or hazards that are associated with the tester.
- g. Wear a beta-gamma type film badge or TLD badge when operating the tester, when in the storage area for the tester, and at other times as directed by the LRPO.
- **4-13. TESTER DISPOSAL.** Notify the NICP item manager and the RMCP when testers are unserviceable, or are in excess of the amount required by the unit.

WARNING

Under no circumstances shall the testers be conveyed to a property disposal officer. All Disposal actions shall be conducted in accordance with AR 385-11.

PERSONNEL DOSIMETRY

- **4-14. RADIACMETERS.** All personnel who enter radioactive storage areas or operate the tester are authorized to wear IM-9E/PD Radiacmeters, NSN 6665-00-243-8199, if required by the LRPO.
- 4-15. FILM BADGES or TLD BADGES. All per

sonnel who enter radioactive storage areas or operate the tester, or at other times, as directed by the LRPO, shall wear beta-gamma type film badges or TLD badges. The Surgeon General has determined that neutron dosimetry badges need not be worn by personnel using the tester.

- **4-16. MEDICAL RECORDS.** The responsibility for preparing and maintaining DA Form 3484 (Photodosimetry Report) or a computer equivalent form (Dosimeter Issue Listing) supplied by the U.S. Army lonizing Radiation Dosimetry Center, is a function of each unit which uses the tester.
- a. The person responsible for maintaining personnel dosimetry records shall be so designated in writing (AR 40-14) as the custodian of medical records. The custodian shall maintain DD Form 1141 (Record of Occupational Exposure to Ionizaton Radiation), and DD Form 1952 (Dosimeter Application and Record of Occupational Radiation Exposure).
- b. Personnel can obtain the total amount of radiation to which they have been exposed, and other related data, by contacting either their LRPO, the custodian of medical records, or the Chief, U.S. Army Ionizing Radiation Dosimetry Center, ATTN: AMXTM-S-LR-DN, Lexington, KY 40511-5102.

STORAGE

4-17. UNOCCUPIED AREAS

- a. It is mandatory that the testers be stored in locked, unoccupied and isolated areas. Fire resistant structures which provide protection from the weather shall be used when they are available. Only work that is necessary for normal storage and maintenance procedures shall be performed in these storage areas.
- b. Modified Table of Organization and Equipment (MTOE) authorizes one AN/PDR-27 Radiac Set or equivalent for each work group requiring the protection of radiation monitoring.

- c. An instrument survey, using the AN/PDR-27 or equivalent, shall be performed before the tester is placed in a storage area. This is done to determine background radiation level of the storage area, thus providing a measurement baseline in the event of contamination.
- d. An instrument survey of the area surrounding any new storage facility shall be made immediately after radioactive material has been placed in the storage facility. The survey shall be recorded, and shall contain the information which is detailed in Para4-23.
- e. Maximum dose equivalence rate within the storage area shall not exceed 2.5 millirems per hour (mrem/hr).

4-18. OCCUPIED AREAS

- a. The number of testers that can be safely stored, in either occupied buildings or in occupied areas, is determined by shielding, the distance from the source and that part of the area which is occupied. Maximum dose equivalence rates outside of storage areas that are within occupied buildings and occupied areas shall not exceed 0.25 mrem/hr.
- b. A tester which has been packed in its approved carrying case may be stored in a supply area or weapon storage area, providing the following requirements have been met:
- (1) The cased tester shall be locked in a metal storage cabinet.
- (2) The storage cabinet shall be isolated and located as far as possible from personnel work stations.
- (3) All required radiation caution signs shall be displayed on the storage cabinet and on the door leading into the storage area.

- (4) Floor markings shall be so placed as to identity a 2-meter area around the storage area. The marked area should only be entered by the LRPO or operator.
- c. An instrument survey shall be performed in accordance with Para 4-23.

4-19. ADMINISTRATION

- a. Standard radiation caution signs, as described in AR 385-30 and Title 10, CFR Part 20, shall be conspicuously posted in places where radiation is present. Installations or activities located where non-English languages are prevalent should post signs that include a translation in those languages. A sufficient number of signs shall be posted so that they are clearly visible from all approaches to the area. The following documents shall be posted at the storage area entrance unless indicated below
- (1) NRC Form 3. [Can be obtained from AR 385-11or the RCO).
- (2) Title 10 CFR Parts 19, 20, 21. (Can be obtained from the RCO).
- (3) Section 206, Energy Reorganization Act of 1974 (Public Law 93-438) (Located at the back of this manual.)
- (4) A copy of the NRC License. (Can be obtained from the RCO).
- (5) A copy of radiation protection program SOP. (A sample SOP is located at the back of this manual).

Copies of these documents can be obtained from the Commander, TACOM, ATTN: AMSTA-CZ. Posting of Title 10 CFR Parts 19, 20, 21, NRC license and the radiation protection program SOP is not required when a notice is posted at the storage entrance as to where the documents can be reviewed.

b. Storage areas shall be classified as restricted areas.

NOTE

Individual testers shall be labeled with Department of Transportation (DOT) Yellow Label II shipping labels, and other required labeling in accordance with AR 385-30.

- c. The local tire department shall be notified of radioactive storage locations, and where to contact the LRPO. The installation LRPO shall also be informed of the storage locations.
- d. The time that personnel spend in storage areas shall be kept to an absolute minimum.
- e. An inventory shall be maintained with a physical inventory count conducted semiannually. The inventory shall record the National Stock Number (NSN), Serial Number (SN), Special Item Control Code (SICC), radioisotope(s) in device, original activity of radioisotope(s), chemical and physical form, storage location, number of items and the license or authorization number it is under.
- f. Providing that the requirements of Para 4-17, 4-18 and 4-19 have been complied with, any number of testers can be stored in an approved storage area.

TRANSPORTATION OF TESTERS BY USING UNITS

4-20. GENERAL

a. A letter of instruction shall accompany the tester when it is moved from a government controlled facility, and from one job site to another. A sample letter of instruction is located in Appendix G. DD Form 836 (Special Instructions For Motor Vehicle Drivers), may be used as the letter of instruction. The general requirements of AR 385-11 shall be complied with, as shall State and Local requirements. It is mandatory that the tester be packed and locked in its approved carrying case during transport. The case shall be properly secured to prevent damage to it and the tester during transport.

- b. Testers to be shipped or moved around, shall meet the requirements of Para 5-16 (Shipping Instructions).
- **4-21. AIR DROP.** The tester is not authorized for air drop.

4-22. AIR TRANSPORTATION

- a. The tester is not authorized for shipment on commercial passenger aircraft. Commercial air transportation is governed by DOT regulations.
- b. The tester is approved for air transport by helicopter, fixed-wing Army aircraft, Air Force aircraft, and commercial nonpassenger-carrying aircraft when shipped in accordance with TM 38250, AR 385-11, and applicable DOT and NRC regulations.

SURVEYS AND TESTS

- **4-23. RADIATION SURVEYS.** Radiation surveys shall be conducted at least quarterly. Records of surveys shall be maintained by the Radiological Protection Officer. A copy of the initial survey, while the tester is in the storage area, shall be provided to Commander, TACOM, ATTN: AMSTACZ. A closeout radiation survey shall be made and documented for the tester storage location when the storage of the tester is terminated at that location. A copy of this survey shall be provided to the TACOM Safety Office, AMSTA-CZ. The surveys shall contain, as a minimum, the following information:
- a. A sketch of the storage area showing tester location and adjacent offices and buildings.
- b. Recorded radiation levels and location of the survey readings.
 - c. Distance from the tester to the survey location.
 - d. Location of radiation caution signs.
- e. Identification of the instrument used to make the survey, and the date on which the instrument was calibrated.

- f. Name of the surveyor.
- g. Date of the survey.
- h. A statement attesting to the fact the documents specified in Para 4-19a have been posted.
- i. Any corrective actions required and accomplished.

4-24. MINIMUM REQUIREMENTS FOR SCHEDULED SURVEYS

- a. Check the security of radiation sources and testers.
- b. Ensure that the area is posted in accordance with AR 385-30, Title 10 CFR Part 20 and Para 419a.
- c. Ensure that personnel comply with the radiation safety procedures.
- d. Ensure that personnel comply with the SOP for safe handling of the tester and tester security.
- e. When a tester storage area is relocated or renovated, perform a new survey in accordance with Para 4-23. Furnish a copy of the new survey to Commander, TACOM, ATTN: AMSTA-CZ.

4-25. RADIATION LEAK (WIPE) TEST

- a. In compliance with AR 385-11, every tester shall be tested for radiation leakage at intervals not to exceed six months. The absence of a certificate of transfer, which would have indicated that a leak test had been made within six months prior to transferring a tester, mandates that the tester shall not be used until it has been tested. The leak test requirement applies to all testers, both in use and in storage. See Para 4-26 for leak test procedures.
- b. Tester users in the U. S.,Puerto Rico,Panama and Korea shall receive leak test kits every six months, January and June. Units in these areas not receiving the kits during this timeframe shall notify Commander, TACOM, ATTN: AMSTA-CZ.
 - c. Leak test results shall be maintained by the

LRPO, and shall be recorded in microcuries.

- d. Wipe tests made by users in the U.S., Puerto Rico, Panama and Korea shall be sent to: Chief, us Army Ionizing Radiation Dosimetry Center, AMXTM-S-LR-DN, Lexington, KY 40511-5102.
- e. Wipe tests made by USAREUR units shall be sent to: Radiac ACL, Pirmasens, 524th Maintenance (TMDE) Co., APO NY 09138-4628.
- f. Overseas users in other than USAREUR units will be notified of Army laboratories to which wipe tests shall be sent.
- g. Units that receive leak test kits every six months are not required to send test results to TACOM. The test results are forwarded to TACOM by the U.S. Army lonizing Radiation Dosimetry Center.
- h. Users who do not receive leak test kits every six months, shall send test results to Commander, TACOM, ATTN: AMSTA-CZ. Test results shall include the LRPO's name, and a statement that the tester is being used only by trained personnel.
- (1) The LR PO shall ensure that test results are recorded and submitted on DA Form 3252-R (Punch Card Transmission Worksheet Radioisotope Inventory and Leak Test Report), as required by AR 385-11 and Para 4-9e.
- (2) When filling out the DA Form 3252-R, use the last five digits of the tester serial number for blocks 18-22.
- i. The LRPO shall immediately withdraw a tester from use when results of leak testing reveal the presence of 0.005 or more microcuries. Immediately thereafter, the LRPO shall notify the installation's RPO, the RCO, and the Commander, TACOM. The telephone shall be used to notify TACOM. A request for instructions from TACOM shall also be made during the telephone call. TACOM (AMSTA-CZ) telephone numbers are:
- (1) During office hours: DSN 7866121/6194, Commercial (313) 574-6121/6194.
- (2) After office hours: DSN 786-5511, Commercial (313) 574-5511.

4-26. LEAK TEST PROCEDURE

- a. The tester contains two sealed sources. Both of the sources must be tested for radiation leakage to comply with AR 385-11. See Para 4-25.
- b. The following items are required to performing the leak test:
 - (1) Transparent plastic wrapping material.
 - (2) Pressure sensitive transparent tape.
 - (3) Cotton swab with wooden stem.
 - (4) Detergent and water solution.
 - Aluminum or other metallic foil.
 - (6) Radiac meter AN-PDR/27 or equivalent.

4-27. LEAK TEST

- a. Place a sheet of clean plastic wrap on an uncontaminated surface near the tester.
- b. Make sure the testers handle is locked in the "SAFE" position, and that the shutter is closed. (See Fig. 4-1).




Figure 4-1. Model MC-1 Nuclear Method Density and Soil Tester

- c. Carefully place the tester on its end or side. (See Fig. 4-2).
- d. Moisten the cotton swab in the detergent/water solution.
- e. Swab the ring which is located on the bottom of the tester. Do not swab the source rod or the shutter. (See Fig. 4-2).
- f. Place the cotton swab on the sheet of plastic wrap.
- g. Return the tester to its upright position. (See Fig. 4-1).




Figure 4-2. Swabbing the Ring

- h. Loosen the four screws which are located in the corners of the tester's display panel. Lift the panel out of the tester and support it on the case. (See Fig. 4-3).
- i. With the same swab that was used in e above, swab the red spot which is located within the tester and next to the internal radiation detector. The spot to be swabbed may not be painted red in some testers. (See Fig. 4-3).
- j. Wrap just the cotton swab and an inch or two of the stick with plastic wrap. Seal the plastic wrap around the stick with cellophane tape.




Figure 4-3. Swabbing the Spot

- k. Reinstall the display panel in the tester and tighten the four screws.
- I. The swab shall be tested for radiation in an area where background radiation averages 0.04 mrem/hr or less.
- (1) Adjust the radiac meter AN-PDR/27 or equivalent to measure 0 to 0.5 mrem/hr.
- (2) Open the cover end of the radiac meter's probe.
- (3) Do not touch the swab to the meter's probe. Place the swab about 0.25 inch in front of the probe and note the meter reading.

WARNING

- A tester is contaminated when a reading of 0 I mrem/hr or more (or in the case of Colorado 0.2 mrem/hr or more) is obtained from the swab.
 Follow the procedures provided in Para 4-31.
 The swab is also contaminated.
- The tester shall immediately be withdrawn from use when the wipe test indicates a reading of 0.1 mrem/hr. (or in the case of Colorado 0.2 mremlhr) or when the laboratory test results reveal the presence of 0.005 or more microcuries.
- m. Cut off the applicator stick at the tape seal.
- n. If a swab reading greater than 0.4 mrem/hr

was obtained, wrap the sealed swab in several layers of aluminum, lead or other metallic foil, and place it in a small cardboard box. Make sure the box is closed.

- 0. Monitor the cardboard box for surface radiation. The reading must be below 0.4 mre/hr before the swab can be mailed to the analysis laboratory. Add additional wrappings of foil to reduce surface
- radiation below 0.4 mrem/hr.
- (1) In the US, Puerto Rico, Panama, and Korea, mail swabs to: Chief US Army Ionizing Radiation Dosimetry Center ATTN: AMXTM-S-LR-DN Lexington, KY 40511-5102
- (2) USAREUR units will forward swabs to: Radiac ACL, Pirmasens, 524th Maintenance (TMDE) Co., APO NY 09138-4628.
- (3) Other tester users shall forward swabs to laboratories indicated in procedures established by the area commander.

4-28. RECORDS

- a. Record the data of laboratory test results in a permanent file.
- b. Units in the US and Puerto Rico that do not receive wipe test kits from the wipe test kit service shall mail a copy of the laboratory test results to:

Commander US Army Tank-Automotive Command ATTN: AMSTA-CZ (Safety Office) Warren, MI 48397-5000

EMERGENCY PROCEDURES 4-29. LEAKS, LOSS OR TESTER DAMAGE (FIRE AND EXPLOSION)

a. To become qualified, all tester users must have had previous training in procedures for handling emergencies that could occur during transportation of the tester.

- b. Personnel responsible for storage facilities must have had previous training in procedures to be followed during emergencies that could occur when the tester is in storage.
 - (1) Panic must be averted during emergencies.
- (2) The first person to discover a radiological incident on a military installation must immediately notify the installation's RPO and Commander, TACOM, ATTN: AMSTA-CZ.
- c. When a tester is exposed to or damaged by fire or explosion, personnel who have been trained in emergency procedures shall perform the following duties, in the order given.
- (1) Notify all personnel who are not directly involved in the emergency to clear the area.
- (2) Notify fire fighting and other emergency personnel of the situation.
- (3) When a radiological hazard is not immediately imminent, attempt to extinguish fires with firstaid type extinguishers. Try to prevent water or firefighting chemicals from coming into contact with radioactive sources.
 - (4) Notify the LRPO of the situation.
 - d. After the emergency, the LRPC shall:
- (1) Monitor all personnel who were in the emergency area.
- (2) Monitor all personnel who were involved in combating the emergency.
- (3) Monitor the area and determine those protective measures that are necessary for the removal of radiation hazards.
- **4-30. PHYSICAL LOSS OF TESTER.** When a tester is lost, the LRPO shall attempt to recover the tester by:
 - a. Making a physical search for the tester.
 - b. Reviewing records to determine the individual

who was responsible for the tester.

- c. When the tester cannot be recovered, the LRPO shall report the loss to the Radiation Material Control Point (RMCP) and the Commander, TACOM, ATTN: AMSTA-CZ. The report shall include: (1) Serial number of the lost tester.
 - (2) Circumstances concerning the loss.
- (3) Action taken to prevent recurrence of tester loss.
- d. When the tester is recovered, revise procedures to prevent recurrence of tester loss.

4-31.DAMAGED OR LEAKING TESTER

- a. When a tester is leaking radiation, or is suspected of leaking, the LRPO shall: (1) Immediately discontinue use of the tester.
- (2) Wrap the tester in plastic film and seal the package with tape.
- (3) Conspicuously label the package as being contaminated with radioactivity.
- (4) Monitor personnel, equipment, and the area for contamination. Decontaminate as required.
- (5) Report that the tester is damaged, or leaking, to both the RMCP and the Commander, TACOM, ATTN: AMSTA-CZ.
- (6) Dispose of the tester as directed by the NICP, and comply with AR 385-11.
- (7) Report the completed disposal action to both the RMCP and the NICP.
- b. Shutter Failure. When the tester's source shutter fails to close or is damaged, the LRPO shall:
- (1) Immediately discontinue use of tester. Do TM 5-6635-386-12&P

not attempt to repair the tester.

- (2) Place tester in shipping case and lock the case.
- (3) Place shipping case in an isolated, locked area and secure it against unauthorized handling.
- (4) Report immediately, by telephone, to both the RMCP and Commander, TACOM, AMSTA-CZ, for further instructions.

4-32. PERSONNEL EXPOSED TO RADIATION

- a. Overexposure.
- (1) The individual(s) shall immediately report to the medical officer in the event of suspected or actual overexposure.
- (2) The individual(s) shall notify the LRPO of the suspected or actual overdose.
 - b. Skin.
- (1) An individual's skin is considered overexposed when it monitors at a higher instrument reading than that of the previously determined background level.
- (2) The individual shall immediately wash, paying particular attention to cleaning fingernails, the hair, and creases in the skin.
- c. Cuts and Skin Abrasion Thoroughly flush the wound with water when contamination is either suspected or confirmed.
 - d. Ingestion.
- (1) Do not smoke, eat or drink when handling the tester or any other radioactive Item.
- (2) Immediately inform the medical officer when ingestion of radioactive contamination is suspected.
 - e. Clothing.
 - (1) Immediately remove clothing that monitors

- at 0.2 mrem/hr or more at a distance of 2.5 centimeters (about one inch).
- (2) Place clothing in a container. Seal and label the container as radioactive waste. A garbage can with attached lid, and a plastic bag that has been sealed shut with tape are both considered to be sealed containers.
- f. Rags. Place all clean-up rags in sealable containers.
- g. Disposing of radioactive waste. Comply with AR 385-11 when disposing radioactive waste.
- **4-33. PERSONNEL EXPOSED TO EXCESSIVE RADIATION**. The following actions shall be taken should an individual receive a dose of ionizing radiation from the tester which exceeds 1.25 rem per calendar quarter, or an accumulated dose in excess of 5 rem per calendar year. Refer to Para 5-30, AR 40-5, for procedures for personnel exposed to excessive radiation.
- a. The individual shall immediately seek advice from the medical officer, and then notify the unit commander of the overexposure.
- b. The unit commander shall remove the individual from duties involving occupation exposure to radiation until subsequent exposure limitations are established by proper medical authority (AR 40-14).
- c. The unit commander, with assistance from the LRPO, shall prepare and submit a written report of circumstances leading to the overexposure. The report shall include the following.
- (1) Serial number of the tester that caused the overexposure.
- (2) Action taken to prevent recurrence of overexposure.
- d. Overexposure reports shall be forwarded, through command channels, to the Commander, TACOM, ATTN: AMSTA-CZ.

4-34. DECONTAMINATION OF EQUIPMENT AND AREA

- a. If the radioactive source is determined to be leaking or broken open, it may contaminate surroundings requiring decontamination efforts. Make sure not to spread the contaminant during the decontamination process.
- (1) Ensure that contamination has not spread by monitoring the area of contamination with survey instrumentation while wearing protective clothing and shoe covers.
- (2) During the decontaminating process, always work from the areas of least contamination toward the areas of greatest contamination.
- (3) Use the minimum amount of decontaminating solutions. Remain alert to the fact that run-off solutions, mops, rags and brushes are contaminated.
- b. Accomplish decontaminating methods in the following order of priority.
- (1) Damp mopping. Wipe the contaminated area with a damp rag or a damp mop. Keep changing the wiping surface of the rag or mop to minimize spreading the contaminant.
- (2) Water and detergent. Wet the contaminated area with a rag or mop which has been dampened with a minimum amount of detergent solution. Then, wipe the area dry with absorbent gauze or cloth.

WARNING

Before proceeding to the next decontamination level, contact the Commander, TACOM. ATTN: AMSTA-CZ for guidance concerning respiratory protection. Failure to wear respiratory protection may result in overexposure to radiation.

- (3) Steam cleaning.
- (4) Cleaning with solvents other than water.
- (5) Surface removal by use of chemicals, abrasives, sand blasting and grinding.

- (6) The only vacuum cleaners that shall be used are those with absolute filters which have been tested for filtration efficiency. Testing filtration efficiency is mandatory each time the filter is replaced, and each time its contents are emptied.
- c. Radioactive waste shall be disposed of in a manner to comply with AR 385-11.

MAINTENANCE

4-35. TESTER MAINTENANCE

- a. Only limited and specific maintenance is authorized for the tester. Personnel authorized to perform maintenance are the operator and the LRPO. Neither the operator nor the LRPO are authorized to repair the tester (see Para 4-37).
- b. Operator performed maintenance is limited to the following:
- (1) Cleaning the testers external surfaces and shutter assembly.
 - (2) Changing fuses.
 - (3) Changing batteries.
 - (4) Charging batteries.
- (5) Perform leak test in absence of qualified LRPO or ALRPO.
- c. LRPO performed maintenance is limited to the following:
 - (1) Perform the leak test.
- (2) When required, the LRPO shall perform maintenance that is normally accomplished by the operator.

4-36. TESTER REPAIR AND DISPOSAL

a. A tester requiring repair or disposal must be returned to a designated manufacturer's repair facility.

- b. Do not use the tester when it fails to operate in accordance with preoperation checks described in the operator's manual.
- c. Contact the NICP for disposition instructions. See Para 4-38a.
- d. Contact the Commander, TACOM, ATTN: AMSTA-MVC if all operating modes of the tester are not functioning correctly. Users are requested to submit a Quality Deficiency Report (QDR) on Standard Form 368 in accordance with DA Pam 738-750 to the address in Para 4-39b.

ASSISTANCE

4-37. SAFETY. Copies of surveys, leak test results, incident reports, other applicable data and reports, DA Form 3252-R, and questions pertaining to the tester and personnel safety shall be directed to:

Commander

US Army Tank-Automotive Command ATTN: AMSTA-CZ (Safety Office)

Warren, MI 48397-5000

Telephone: Autovon:

786-6121 or 61 94

Commercial: (313) 574-6121 or 6194

After office hours: Autovon: 786-5511

Commercial: (313) 574-5511

4-38. TESTER REQUISITION, AND DISPOSITION.

Questions or requests pertaining to tester requisition, and disposition shall be directed to:

a. Commander

US Army Tank-Automotive Command

ATTN: AMSTA-FHV (NICP) Warren, MI 48397-5000

Telephone: Autovon: 786-5827 or 7787 Commercial: (313) 574-5827 or 7787

Questions pertaining to tester maintenance shall be directed to:

b. Commander

US Army Tank-Automotive Command ATTN: AMSTA-MVC (Maintenance)

Warren, MI 48397-5000

Telephone: Autovon: 786-8298/7358 Commercial: (313) 574-8298/7358

4-39. TRAINING COURSES. Questions pertaining to training courses shall be directed to:

a. Material Quality Specialist course (51 $\,$ G) - including tester training.

Commander HO, 87th Eng Bn

ATTN: 51 Committee (51G) Ft. Leonardwood, MO 65473-6400 Telephone: Autovon: 581-3453 or Commercial: (314) 596-3453

b. Calibrator Custodian Course.

Commander TRADOC

ATTN: ATTG-MPS

Ft. Monroe, VA 23651-5000 Telephone: Autovon 680-2161 or Commercial (703) 664-2889

c. Radiological Safety Course 7K-F3.

Commander TRADOC

ATTN: ATTG-MPS

Ft. Monroe, VA 23651-5000 Telephone: Autovon 680-2161 or Commercial (703) 664-2889

d. Army National Guard Radiation Protection Officers Course.

Commander CECOM

ATTN: AMSEL-SF

Ft. Monmouth, NJ 07703-5000 Telephone: Autovon 995-4427 or Commercial (201) 544-4427

e. Operational Radiation Safety Course 4J-F2

Commander TRADOC

ATTN: ATTG-MP5

Fort Monroe, VA 23651-5000 Telephone: Autovon 680-2161 or Commercial (703) 664-2889

SECTION V

SUPPLEMENTAL OPERATING, MAINTENANCE AND REPAIR PARTS INSTRUCTIONS (SOMARPI) FOR TESTER, DENSITY AND MOISTURE (SOIL AND ASPHALT) NUCLEAR METHOD

CAMPBELL PACIFIC NUCLEAR CORPORATION MODEL MC-1 (NSN 6635-01-030-6896)

5-1. INSTRUCTIONS

- a. These instructions contain supplemental information for Army users and are Army oriented instructions to assist in managing and expediting maintenance repair parts supply, technical assistance, and user problems.
- b. Weapons System Designator Code for requisitioning is 3M.

5-2. GENERAL

- a. The Density and Moisture Tester is a Commercial Construction Equipment (CCE) type used to measure the density and moisture content of soils, soilstone aggregates, and also the density of hot asphalt while the mix is being compacted.
- b. Operation of the tester will be only by personnel who have successfully completed an Operators Training Course as defined in Section IV. Using units and operators shall not operate testers without an assigned qualified LRPO. Operator's will be supervised by a Local Radiation Protection Officer (LRPO) who has successfully completed the training requirements as stated in Para 4-9.c. Radiation Control Officers of Major Commands will monitor the LRPO's or operators (in the case of a using unit without qualified LRPO or ALRPO). The TACOM Safety Office will insure compliance with the Nuclear Regulatory Commission (NRC) license by monitoring reports of inventory, leak tests, incident and equipment improvement reports, by utilizing feedback from Logistics Assistance and Inspection Teams, and by liaison with Radioactive Material Control Points (For details, see Section IV).

5-3. MAINTENANCE CONCEPT

- a. Organizational maintenance consisting of external cleaning, battery charging and replacement, fuse changing, shutter cleaning/lubricating, will be performed by the operator. Wipe (leak) tests will be performed by the LRPO or Operator (in the absence of a qualified LRPO or ALRPO at the using unit).
- b. TACOM, as required, will issue a Task Order/Work Directive/Order for all supplies and services for major repairs to the contractor authorizing necessary work. TACOM can place a dollar limit on work authorized. TACOM will furnish users shipping instructions of designated repair facility. (Users will retain and use the original container, for shipping testers for repair).
- c. Contractor will repair or overhaul testers and ship them back to users.
- d. The contractor will bill the government for his labor, materials, parts and shipping cost.
- e. TACOM will obtain confirmation from the using unit that repairs have been accomplished and will authorize payment of the repair bill.
- f. The contractor will be responsible for disposing of unserviceable, unrepairable testers upon receipt of disposal instructions from NICP at TACOM.

5-4. SUPPORT ORGANIZATIONS

- a. Maintenance None. All maintenance above organization (see Para 5-3.a.) will be obtained by evacuation to the designated contractor's facility.
- b. Supply DS units backing up the using units will process requests for parts shown on the PLL

and ASL. See Appendix D.

5-5. REPAIR PARTS AND SUPPORT. Requisitions will be forwarded to and filled by DCSC. Routing identifier code (S9C) under milstrip procedures (AR 725-50) for non-NSN items see Appendix H, for NSN Items see Appendix 1. Requisitions should cite manufacturer's part numbers in the absence of NSN's.

5-6. PERSONNEL AND TRAINING

- a. All operational and support personnel will require training on the safe handling of nuclear devices (see Para 4-12).
- b. MOS Training: TRADOC will manage and coordinate the MOS training of operators. This training will be conducted by the 51 Committee (51 G) at Ft. Leonardwood, MO.

5-7. EQUIPMENT PUBLICATIONS

- a. Initially two sets of the Commercial Instruction Manuals, supplemented by a separate SOMARPI and TB 385-1 03 (Safety Procedures) were issued to support the soil tester.
- b. After initial issue of Tester, the manufacturers manuals, the SOMARPI and TB 385-103 were combined into this manual.

5-8. FACILITIES

- a. No special maintenance facilities are required for the tester.
- b. For storage facilities, see Section IV, Para 4-17 and 4-18.

5-9. EQUIPMENT IMPROVEMENT

RECOMMENDATION (EIR'S). EIR's will be submitted in accordance with instructions contained in DA PAM 738-750.

- **5-10. DESTRUCTION TO PREVENT ENEMY USE.** When the tester cannot be evacuated the following will be done:
- a. Remove the circuit boards and display panel (loosen and remove the four screws located in the

corners of the tester's display panel, pull off panel, and cut or disconnect wiring).

- b. Destroy the display panel and circuit boards by crushing.
- c. Destroy worksheet and logs by crushing, tearing, and burying, or scattering over a sufficiently large area to make recovery improbable.
 - d. Abandon the remainder of the tester.
- **5-11. PROCEDURE**. In addition to operating and maintenance instructions, see Section IV, Safety Procedures, in this manual.
- **5-12. MANUFACTURER'S SAFETY DEFECTS.** If any manufacturer's safety defects occur, upon notification TACOM will advise users on action to be taken for correction and reporting.
- **5-13. MANUFACTURER'S FIELD CAMPAIGNS AND MODIFICATIONS.** Modifications are not anticipated. If necessary, they will be applied by CPN after TACOM's approval of the field campaign or modification plan.
- **5-14. BASIC ISSUE ITEM LIST.** List of items which accompany the tester or are required for installation, operation. or operator maintenance (Appendix C).
- **5-15. PERSONNEL DOSIMETRY.** See Section IV in this manual.
- **5-16. SHIPPING INSTRUCTIONS**. Shipping Information for Tester, Density and Moisture, Nuclear Method, NSN 6635-01-030-6896. For additional information see paragraph 4-20.
- a. Consult with the unit Transportation Officer to insure that the tester is shipped in accordance with Department of Transportation Regulations (Title 49 CFR Parts 170-189) AR 385-11, TM 38-250, and the Nuclear Regulatory Commission Regulations (Title 10 CFR Part 71). For CONUS, a commercial shipper should be used to transport the tester.
- b. The following information is provided to assist in the preparation of shipping papers for the tester (Title 49 CFR Parts 172-173):

- (1) Proper Shipping Name: Radioactive Material, Special Form, N.O.S.
 - (2) Hazard Class: Radioactive Material.
 - (3) Identification Number: UN 2974.
 - (4) Total Quantity: 1 Box.
- (5) Name of each Radionuclide: Americium241, Cesium-137.
 - (6) Form: Special Form.

- (7) Activity: 60 millicuries total, 10 millicuries Cesium-137, 50 millicuries Americium-241 (8) Label Required: Radioactive Yellow-I.
- (9) Transport Index: Use value listed on Yellow II label on box.
- (10) The following accessories are to be included with the testers when transporting them for repair/overhaul or to a new location:

DESCRIPTION	FSCM	PART NO.	NSN
Cast Aluminum Guideplate	55492	B 200050	6635-01-043-7970
Computer Tabulations AC Charger	55492	NB-150-8	6635-01-043-7972
DC Charger (Battery Clips)	55492	A 401234	6635-01-043-7973
Sign Set	55492	W1006	9905-01-115-8157
Phillips Screwdriver	79061	P28	5120-00-542-3438
Fiberglass Case	55492	A 401228	No NSN
Padlock w/2 Keys	96906	MS 35647-4	5340-00-582-2741
Standard Drill Pin	55492	A200128	6635-01-043-7971
Clipboard Assembly	55492	B 101013	7520-01-115-8066
Standard Block Assembly	55492	B 401124	No NSN
Wrench, Allen Hex, 5/32	80064	1940717	5120-00-198-5392
Battery, NICAD, 8 ea	19209	41B004AA71WTABS	6140-00-497-0490
(Size D-4 Ampere hour 160F)			

^{*}Some testers were issued with a hammer. If applicable, the hammer should also be included. TM 5-6635-386-12&P shall always stay with the tester even during shipment.

- (11) Handle must be locked in a safe position and the batteries removed before tester is shipped.
- (12) The key for the padlock is to be tagged with the tester serial number and sent by registered mail to the receiving/repair facility. The key to the probe handle shall be stored inside the case.
- c. Additional information is required as appropriate depending on the mode of transportation used. Do not forget to include the following shipping certification on the shipping papers:

This is to certify that the above named materials are properly classified, described, packaged, marked and labeled and are in proper condition for transportation according to the applicable regulations of the Department of Transportation.

- d. The orange tester case meets the requirements of specification 7A for type A packaging. Type A packaging is required for the tester since the radionuclides are in special form and the amount of activity of the two radioactive sources is below the maximum permitted for type A packaging (Al). However, to protect from damage during shipment, the tester and case should be placed in a wooden crate. The outside of the crate must be labelled and marked in accordance with Title 10 CFR Parts 172.310, 172.403 and 438. A complete DD Form 1348-1 (DOD Sirtgle Line Item Release/Receipt Document) should accompany the shipment to the manufacturer.
- e. The Transport Index (TI) must be marked on the Dot Radiation label on the tester package, and on the shipping paper. The TI is a number expressing the maximum radiation level in millirem per

hour at one meter (3.3 feet) from the external surface of the package. The number must be rounded up to the first decimal place. For example, a reading of 0.11 should be rounded to 0.2. The typical TI number for testers is .05. The TI is used to designate the degree of control to be exercised by the carrier during transportation.

- f. When a tester shipment is received, immediately perform a wipe test to determine whether or not the radioactive sources were damaged during shipping.
- g. Only serial numbered testers listed on item manager letter for repair/overhaul should be sent to the manufacturer. No replacements are allowed without notification of item manager, AMSTA-FHV, AV 786-7787/5827 or Commercial (313) 574-7787/5827. (See Para 4-20 for additional information).

5-17. OPERATOR/UNIT PREVENTIVE MAINTENANCE CHECKS AND SERVICES (PMCS)

NOTE

Your Preventive Maintenance Check and Services table lists the inspection and care of your equipment required to keep it in good operating condition.

- a. The number column of your PMCS is the source for the number used on the TM column on DA Form 2404.
- b. The interval column of your PMCS table tells you when to do a certain check or service.
- (1) Before you operate. Always keep in mind the WARNING S and CAUTIONS. Perform your before (B) PMCS.
- (2) While you operate. Always keep in mind the WARNINGS and CAUTIONS. Perform your during (D) PMCS.

- (3) After you operate. Be sure to perform your after (A) PMCS.
- c. The procedure column of your PMCS table tells you how to do the required checks and services. Carefully follow these instructions. If you do not have the tools, or if the procedure tells you to, contact the unit maintenance.
- d. If your equipment does not perform as required, refer to the manual troubleshooting section for possible problems. Report any malfunctions or failures on the proper DA Form 2404 or refer to DA Pamphlet 738-750.

NOTE

The terms ready/available and mission capable refer to the same status: Equipment is on hand and is able to perform all its combat missions without further endangering the lives of crew or operators in a combat environment (see DA Pamphlet 738-750).

- f. Always do your PMCS in the same order so it gets to be a habit. Once you've had some practice, you'll spot anything wrong in a hurry.
- g. While performing PMCS, observe WARNING and CAUTIONS preceding those operations which could endanger your safety or result in damage to the equipment.
- (1) Keep it clean; dirt, grease, oil and debris only get in the way and may cover up a serious problem. Use dry cleaning solvent (P-D-680) to clean metal surfaces.
- (2) Bolts, nuts, and screws: check that they are not loose, missing, bent or broken. Tighten any bolt, nut, or screw that you find loose.

Table 5-1. Operator's Preventive Maintenance Checks and Services

		Location		
Item No.	Interval	Item to Check/ Service	Procedure	Not Fully Mission Capable If:
	tester and up Protection Of The tester sh mrem/hr or m twice the bac	on entering tester storage ficer (LRPO), Beta-gamm nall immediately be without ore, except in the case of	warning or thermoluminescent dosimetry TLD badge areas, and at other times as directed by the a badge or TLD badge will be exchanged every warning. Warning Brawn from use when the wipe test indicate of the state of Colorado in which wipe test rewhen the analysis laboratory test result rev	Local Radiation ery 30 days. es a reading of 0.1 eadings at or above
	in which read		warning of 0.1 mrem/hr or more, except in the case of	
1	Before	Batteries	Check the batteries and holder for leakage and, ensure that only NICAD batteries are being used.	Batteries are leaking
			5-5	I

Table 5-1 Operator's Preventive Maintenance Checks and Services

Item No.	Interval	Item to Check/ Service	Procedure	Not Fully Mission Capable If:
2	During	End Item Batteries	Diagnostic Inspection, check the following: a. Does the display work on all channels? b. Does the tester beep during counting? c. Does the tester count in TEST' MODE? What does it indicate upon completion? Section 1, Para 1-4. d. Does it count in STANDARD MODE? Is this a normal standard count? What is the spread of the standard count? Refer to Section1, Para 1-7 and 1-8. e. Does it count in the field count conditions? In all time sequences? Check the batteries display light window, if "L" is displayed recharge batteries overnight using charger. Charge NICAD batteries with AC charger or DC leads. (Refer to illustration, Page 5-7).	Display not working. Tester doesn't beep during counting. Tester doesn't count/read-out incorrect. Tester doesn't count/read-out incorrect. Tester not counting in all time sequences "L" or nothing in display light window.
			5-6	

Table 5-1 Operator's Preventive Maintenance Checks and Services

		Location		
Item No.	Interval	Item to Check/ Service	Procedure	Not Fully Mission Capable If:
	and use breath v Dryclea while us	e only in a well ventilated a vapors. Do not use near on ning Solvent is 1 00°F (38 sing cleaning solvent, get	AC CHARGER WARNING xic and flammable. Wear protective goggles and glovarea. Avoid contact with skin, eyes and clothes, and open flame or excessive heat. The flash point for Typ 8°C) and for Type #2 is 138°F (500C). If you become fresh air immediately and get medical aid. If contact th water and get medical aid immediately.	don't e #1 dizzy
4	Weekly	End Item	Tester/Shipping Case exterior will be washed with soap and water if it becomes unusually dirty. Remove asphalt with PD-680.	

Table 5-1 Unit Preventive Maintenance Checks and Services

		Location						
Item No.	Interval	Item to Check/ Service	Procedure	Not Fully Mission Capable If:				
			WARNING					
	WARNING							
	areas, and		e or TLD badge when operating the tester an by the Local Radiation Protection Officer (LR 30 days).					
			WARNING					
	more, exce	ept in the case of the state	drawn from use when the wipe test indicates a of Colorado in which wipe test readings at or aboratory test results reveal the presence of C	above twice the background				
	which read		WARNING ng of 0.1 mrem/hr or more, except in the case e, is obtained from the swab. Follow procedure					
1	Monthly	Badge	Beta-gamma badge or TLD badge will be exchanged every 30 days.	Beta-gamma badge or TLD badge has not been exchanged within the last 30 days.				
			WARNING					
	Do not atte	empt to charge ZINC-CAR	BON batteries. ZINC-CARBON batteries are r	not be be used in the tester.				
	Do not atte	empt to charge ALKALINE	batteries. ALKALINE batteries are not to be u	used in the tester.				
2	Semi- annually	Batteries	Check the batteries and holder for leakage and ensure that only NICAD batteries are being used. Charge NICAD batteries with AC charger or DC leads.	Batteries are leaking				
			5-8					

Table 5-1 Unit Preventive Maintenance Checks and Services

		Location		
tem No.	Interval	Item to Check/ Service	Procedure	Not Fully Mission Capable If:
3	Semi-annually	End Item	Perform leak test in accordance with Section 4, Para 4-25, 4-26 and 4-27.	Leak test has not been performed in the last six months, or tester fails leak test.
			RING	SPOT
			5-9/(5-10 blank)	

APPENDIX A

REFERENCES

A-1. SCOPE

This appendix lists all forms, field manuals, technical manuals and miscellaneous publications referenced in this manual.

A-2. FORMS

Moisture and Density Tester (Nuclear) Utilization Log	
A-3. ARMY REGULATIONS	
Preventive Medicine	
A-4. TECHNICAL MANUALS	
Preparation of Hazardous Materials for Military Air ShipmentsTM 38-250	
A-5. DA PAMPHLETS	
The Army Maintenance Management System (TAMMS)DA PAM 738-750	
A-6. TECHNICAL BULLETINS	
Calibration Requirements of the Maintenance of Army MaterialTB 43-180	
A-7. MISCELLANEOUS DOCUMENTS	
Energy Reorganization Act of 1974, Section 206	

A-1/(A-2 blank)

APPENDIX B

MAINTENANCE ALLOCATION CHART

FOR

TESTER, DENSITY AND MOISTURE (SOIL AND ASPHALT), NUCLEAR METHOD

SECTION I. INTRODUCTION

- **B-1. GENERAL**. This Maintenance Allocation Chart designates responsibility for performance of maintenance repair functions at specified maintenance levels.
 - a. Section I is a general explanation and definition of terms.
- b. Section II shows the maintenance level responsible and estimated work measurement time for specific functions.
- c. Section III lists common tool sets and the special tool, test and support equipment required for each maintenance function shown in Section II.

B-2. EXPLANATION OF COLUMNS IN SECTION II.

- a. Column 1, Group number. Column 1 lists group numbers, the purpose of which is to identify components, assemblies, subassemblies, and modules with the next higher assembly.
- b. Column 2, Component/Assembly. Column 2 contains the noun names of components, assemblies, subassemblies, and modules for which maintenance is authorized.
 - c Column 3, Maintenance Functions. Column 3 lists the functions to be performed on the item listed in Column 2.
- d Column 4, Maintenance Category. Column 4 specifies, by the listing of a "work time" figure in the appropriate subcolumn(s), the lowest level of maintenance authorized to perform the function listed in Column 3. This figure represents active time required to perform that maintenance function at the indicated category of maintenance. The number of manhours specified by the "work time" figure represents the average time required to restore an item to a serviceable condition under typical field operating conditions. This time includes preparation time, troubleshooting time, and quality assurance/quality control time in addition to the time required to perform the specific tasks identified for the maintenance functions authorized In the maintenance allocation chart. The subcolumns are:
 - C Operator/Crew
 - O Organizational
 - F Direct Support
 - H General Support
 - D Depot
- e Column 5, Tools and Equipment. Column 5 specifies by code, those common tool sets (not individual tools) and special tools, test, and support equipment required to perform the designated functions.

APPENDIX B - CONTINUED

B-3. THE MAINTENANCE FUNCTIONS ARE DEFINED AS FOLLOWS:

- a Inspect. To determine the serviceability of an item by comparing its physical, mechanical and/or electrical characteristics with established standards through examination
- b Test. To verify serviceability and detect incipient failure by measuring the mechanical or electrical characteristics of an item and comparing those characteristics with prescribed standards.
- c. Service Operations required periodically to keep an item in proper operating condition, i.e., to clean (decontaminate), to preserve, to drain, to paint, or to replenish fuel, lubricants, hydraulic fluids, or compressed air supplies.
- d. Adjust To maintain, within prescribed limits, by bringing into proper or exact position, or by setting the operating characteristics to specified parameters.
- e. Align. To adjust specified variable elements of an item to bring about optimum or desired performance f. Calibrate. To determine and cause corrections to be made or to be adjusted on instruments or test measuring and diagnostic equipments used in precision measurement. Consists of comparisons of two instruments, one of which is a certified standard of known accuracy, to detect and adjust any discrepancy In the accuracy of the instrument being compared.
- g. Install The act of emplacing, seating, or fixing into position an item, part, or module (component or assembly) in a manner to allow the proper functioning of an equipment or system.
- h. Replace The act of substituting a serviceable like type part, subassembly, or module (component or assembly) for an unserviceable counterpart.
- i. Repair. The application of maintenance services (inspect, test, service, adjust, align, calibrate, or replace) or other maintenance actions (welding, grinding, riveting, straightening, facing, remachining or resurfacing) to restore serviceability to an item by correcting specific damage, fault, malfunction, or failure in a part, subassembly, module (components or assembly), end item, or system.
- j. Overhaul. That maintenance effort (service/action) necessary to restore an item to a completely serviceable/operational condition as prescribed by maintenance standards (i e , DMWR) in appropriate technical publications. Overhaul is normally the highest degree of maintenance performed by the Army. Overhaul does not normally return an item to like new condition.
- k. Rebuild. Consists of those services/actions necessary for the restoration of unserviceable equipment to a like new condition in accordance with original manufacturing standards Rebuild is the highest degree of materiel maintenance applied to Army equipment. The rebuild operation includes the act of returning to zero those age measurements (hours/miles, etc.) considered for classifying Army equipments/components.

APPENDIX B - CONTINUED

MAINTENANCE ALLOCATION CHART FOR TESTER, DENSITY AND MOISTURE (SOIL AND ASPHALT) NUCLEAR METHOD (CCE)

SECTION II. - ASSIGNMENT OF MAINTENANCE FUNCTIONS

(1)	(2)	(3)			(4)			(5)
Group		Maintenance	Maintenance Level*				Tools and	
Number	Component/Assembly	Function	С	0	F	Н	D	Equipment
	Battery	Inspect	0.1					
	,	Test	0.1					
		Service	0.1					
		Replace	0.1					
	Shutter/Cleanout	Service	0.1					
	Plate	Replace					*	
	Fuses	Test	0.1					
		Replace	0.1					
	End Item	Inspect	0.1					
		Test		0.2**				
		Service	0.1					
		Adjust	0.1					
		Calibrate					*	
		Replace					*	
		Repair					*	
		Overhaul					*	
		İ				İ		

SECTION III. - TOOL AND TEST EQUIPMENT REQUIREMENTS

Tool or Test Equipment				
Reference	Maintenance		National/NATO	Tool
Code	Category	Nomenclature	Stock Number	Number

NOTE

Wipe test kits will be supplied by the US Army Ionizing Radiation Dosimetry Center.

B-3/(B-4 blank)

By manufacturer

** This operation performed by the LRPO for the leak test

APPENDIX C

COMPONENTS OF END ITEM AND BASIC ISSUE ITEMS LISTS

SECTION I. INTRODUCTION

C-1. SCOPE

This appendix lists components of end item and basic issue items for the Density and Moisture Tester to help you inventory items required for safe and efficient operation.

C-2. GENERAL

The Components of End Item and Basic Issue Items Lists are divided into the following sections:

- a. Section II. Components of End Item. These items are part of the end item, but are removed and separately packaged for transportation or shipment. As part of the end item, these items must be with the end item whenever it is issued or transferred between property accounts. Illustrations are furnished to assist you in identifying the items.
- b. Section III. Basic Issue Items. These are minimum essential items required to place the Density and Moisture Tester in operation, and to operate it. Although shipped separately packaged, BII must be with the Density and Moisture Tester. The illustrations will assist you with hard-to identify items.
- c. This manual is your authority to request/requisition replacement Bll, based on TOE/MTOE authorization of the end item.

C-3. EXPLANATION OF COLUMNS


The following provides an explanation of columns found in the tabular listings:


- a. Column (1) Illustration Number (Illus Number). This column indicates the number of the illustration in which the item is shown.
- b. Column (2) National Stock Number. Indicates the National stock number assigned to the item and will be used for requisitioning purposes.
- c. Column (3) Description. Indicates the Federal item name and, if required, a minimum description to identify and locate the item. The last line for each item indicates the FSCM (in parentheses) followed by the part number.
- d. Column (4) Unit of Measure (U/M). Indicates the measure used in performing the actual operational/maintenance function. This measure is expressed by a two-character alphabetical abbreviation (e.g., ea, in, pr).
- e. Column (5) Quantity required (Qty rqr). Indicates the quantity of the item authorized to be used with/on the equipment.


SECTION II. COMPONENTS OF END ITEM

(NOT APPLICABLE)

SECTION III. BASIC ISSUE ITEMS









(1) ILLUSTRAION NUMBER	(2) NATIONAL STOCK NUMBER	(3) DESCRIPTION FSCM and PART NUMBER	(4) U/M	(5) QTY RQR
1	6635-01-043-7973	Cable, DC charger, battery clips (55492) A 401234	ea.	1
		Case assembly, MC (55492) A 401228	ea.	1
2	7520-01-115-8066	Clipboard assembly, MC-1 (55492) B 101013	ea.	1
3	6635-01-043-7972		ea.	1

SECTION III. BASIC ISSUE ITEMS - CONTINUED



(1) ILLUSTRAION NUMBER	(2) NATIONAL STOCK NUMBER	(3) DESCRIPTION FSCM and PART NUMBER	(4) U/M	(5) QTY RQR
4	6635-01-043-7971	Drill pin, forged (55492) A-200128	ea.	1
5	6635-01-043-7970	Guide plate, cast aluminum (55492) B 200050	ea.	1
6		Key, for A 401128 handle assembly (55492) BPW-1	ea.	1
7	5340-00-582-2741	Padlock w/2 keys (96906) MS 35647-4	ea.	1
	5120-00-542-3438	Screwdriver, Phillips #4 (79061) P28	ea.	1
8	9905-01-115-8157	Sign, kit, caution (55492) W-1006	ea.	1

SECTION III. BASIC ISSUE ITEMS - CONTINUED



TA 230984

(1) ILLUSTRAION NUMBER	(2) NATIONAL STOCK NUMBER	(3) DESCRIPTION FSCM and PART NUMBER	(4) U/M	(5) QTY RQR
9		Standard block assembly, MC (55492) B 401124	ea.	1
10	5120-00-198-5392	Wrench, Allen hex, 5/32 (80064) 1940717	ea.	1

APPENDIX D

PRESCRIBED LOAD LIST (PLL) AUTHORIZED STOCKAGE LIST (ASL)

END ITEM MAKE MODEL

Tester, Nuclear Soil Density/Moisture CPN CORP SIMUL-TEST M82AB114 MC-1 SIMUL-TEST

MFR PART NO NSN SERIAL NUMBER RANGE DATE

M82AB114 6635-01-030-6896 TO

SMR	NATIONAL	PART	FSCM	PART DESCRIPTION	U/M	QTY OF PARTS REQ'D FOR NO. OF END ITEMS		
CODE						PLL	ASL 5 4 5 6 20	
	STOCK NUMBER	NUMBER				1-5	1-5	6-20
	5920-00-227-6179	F02B250V1-1/2A	81349	Fuse, 1.5 Amp Slo-blo AG (Box of 5)	BX	1	5	10
	6140-00-497-0490	41B004AA71 WTABS	19209	Battery, NICAD Size D, 4 Amperehour 160°F Special Characteristic	Ea	8	10	20

D-1/(D-2 blank)

APPENDIX E

EXPENDABLE/DURABLE SUPPLIES and MATERIALS LIST

SECTION I. INTRODUCTION

E-1. SCOPE

This appendix lists expendable supplies and materials you will need to operate and maintain the tester. This listing is for informational purposes only and is not authority to requisition the listed items. These items are authorized to you by CTA 50-970, Expendable/Durable Items (except Medical, Class V, Repair Parts, and Heraldic Items).

E-2. EXPLANATION OF COLUMNS


- a. Column (1) Item number. This number is assigned to the entry in the listing and is referenced in the narrative instructions to identify the material (e.g., "Use cleaning compound, item 1, Appendix E").
 - b. Column (2) Level. This column identifies the lowest level of maintenance that requires the listed item.

C Operator/Crew

- c. Column (3) National Stock Number. This is the National stock number assigned to the item; use it to request or requisition the item.
- d. Column (4) Description. Indicates the Federal item name, and, if required, a description to identify the item. The last line for each item indicates the CAGE (in parenthesis) followed by the part number.
- e. Column (5) Unit of Measure (U/M). Indicates the measure used in performing the actual maintenance function. This measure is expressed by a two-character alphabetical abbreviation (e.g., ea, pr, in). If the unit of measure differs from the unit of issue, requisition the lowest unit of issue that will satisfy your requirements.

SECTION II. EXPENDABLE/DURABLE SUPPLIES and MATERIALS LIST

(1)	(2)	(3)	(4)	(5)
ITEM NUMBER	LEVEL	NATIONAL STOCK NUMBER	DESCRIPTION	U/M
1	С	6850-00-664-5685	Solvent, drycleaning, type II (81348) PD-680	qt
2	С	9150-00-163-9036	Lubricant, molybdenum - disulphide, spray (25873) D 5330	
3	С	9150-00-111-4252	Lubricant, spray (53450) 229-49 A	



APPENDIX F

MC-1 MAJOR SUB-ASSEMBLIES PARTS

REFERENCE DWG C-401216-A

ITEM NO.	PART NO.	DESCRIPTION
1	41B004AA71 WTABS	Battery - cell, NICAD, D, 4 ah
2	B 401247	Battery holder - assembly (less battery - cells)
3	A 400180	Cover battery access, MC
4	F02B250V1-1/2A	Fuse 1-1/2 SB
5	A 401128	Handle - assembly w/ <eys< td=""></eys<>
NI	T 401280	PC - assembly, battery clip

F-1/(F-2 blank)

APPENDIX G

LETTER OF INSTRUCTION

WARNING

This vehicle is carrying radioactive materials.

Military/Commercial Vehicle Permanent Transport Document

PURPOSE. The document provides information required by Title 49 CFR Part 172.203(d) for the movement of the Tester, Density and Moisture, MC-1, by military/commercial vehicles. (Document must be maintained in operator's compartment of vehicle).

Required information:

- 1. Item being Transported: Tester, Density and Moisture, Nuclear Method, NSN 6635-01-030-6896.
- 2. Radioactive Material Special Form N.O.S.
- 3. Identification Number: UN2974.
- 4. 10 mci Cs 137 & 50 mci Am 241/Be.
- 5. Special Form. (Sealed in double encapsulated stainless steel).
- 6. Radioactive-Yellow II.
- 7. Transport Index: (Numeric value as listed on the Yellow II label which is determined by the LRPO, each time the tester is shipped out).
- 8. Number of testers being shipped:
- 9. Unit issued the tester: Mailing address.
- 10. Tester being transported to:

Tester being transported from:


- 11. Date(s) of Travel:
- 12. Certification and supporting safety analysis for special form source, Tester, Density, and Moisture, MC-1, NSN 6635-01-030-6896, LIN W02673, is maintained by the U.S. Army Tank-Automotive Command (TACOM), ATTN: AMSTA-CZ, Warren, MI 48397-5000, Commercial Telephone: (313) 574-7635/6194 or AUTOVON 786-7635/6194.
- 13. Certification and supporting safety analysis for specification 7A packaging TYPE A container for Tester, Density and Moisture, MC-1, NSN 6635-01-030-6896, LIN W02673, is maintained by TACOM, ATTN: AMSTA-CZ, Warren, MI 48397-5000. Commercial Telephone (313) 574-7635/6121 or AUTOVON 786-7635/6121

Additional Requirement:

- 1. Vehicle Placarding not required.
- 2. MC-1 must be secured to prevent shifting or movement during transport.
- 3. MC-1 will be secured within the transport vehicle to provide security measures adequate to prevent unauthorized removal.
- 4. MC-1 must be transported in its original carrying case and the case locked.
- 5. Establish suitable exclusion areas around the shipment and prevent unauthorized entry.
- 6. When undamaged, the packages are safe to handle for short periods. In case of an accident or fire, notify immediately the following personnel:
 - a. Name, Unit Radiation Protection Officer (RPO), phone number.
 - b. Name, Unit Alternate Radiation Protection Officer (ARPO), Phone Number.
 - c. Commander's Name, Commander unit, Phone Number.
 - d. RMCP name and phone number.
 - e. Ms Karen Lapajenko, TACOM RPO Commercial (313) 574-7635/6121, DSN 786-7635/6121 (During Duty Hours) or (After Duty Hours) Commercial 574-5935 or DSN 786-5935. Refer to Section IV, Emergency Procedures for emergency instructions, Para 4-29.

///////CERTIFICATION// The signature of the individual(s) below certify that a verbal briefing cor	
was receivedfrom the unit RPO. (date)	
(Driver signature)	(Date)
(Unit RPO or ARPO signature)	(Date)


APPENDIX H SAMPLE FORMAT- MILSTRIP REQUISITION FOR CCE (NON-NSN)



CARD <u>COLUMN</u>	DESCRIPTION <u>OF DATA</u>	MANDATORY ENTRY <u>FOR CCE</u>
1-3	Document identifier Code	AØB - CONUS AØ2 - Overseas
4-6 7 8-22	Routing Identifier Code Media/Status Code FSCM and Part Number	Always S9C
23-24	Unit of Issue	
25-29 30-43	Quantity Document Number	
44 45-50	Demand Code Supplementary Address	
51 52-53	Signal Code Fund Code	
54-56	Distribution Code CC-54	"F" for CONUS; see AR 725-50 for OCONUS
57-59	CC-55-56 Project Code	Weapon System Code "BWG" for CONUS;
60-61	Priority Code	"JZC" for OCONUS
62-64 65-66	Required Delivery Date Advice Code	

H-1/(H-2 blank)

APPENDIX I SAMPLE FORMAT - MILSTRIP REQUISITION FOR CCE (NSN)



CARD <u>COLUMN</u>	DESCRIPTION <u>OF DATA</u>	MANDATORY ENTRY <u>FOR CCE</u>
1-3	Document Identifier Code	AØA - CONUS AØ1 - Overseas
4-6	Routing Identifier Code	
7	Media/Status Code	
8-22	NSN	
23-24	Unit of Issue	
25-29	Quantity	
30-43	Document Number	
44	Demand Code	
45-50	Supplementary Address	
51	Signal Code	
52-53	Fund Code	
54-56	Distribution Code CC-54	"F" for CONUS; see AR 725-50 for OCONUS
	CC-55-56	Weapon System Code
57-59	Project Code	"BWG" for CONUS;
60-61	Priority Code	"JZC" for OCONUS
62-64	Required Delivery Date	
65-66	Advice Code	

I-1/(I-2 blank)

9	2	0	• 7		• 5		7
10		4		• 6		3	
	• 5		3		• 4	• 1	9
•		8	6		10	8	2

APPENDIX J

AREA CONCEPT TESTING


AREA CONCEPT TESTING provides for the evaluation of a site by use of several readings on the site to give a larger representative sample than a single test would give. The average of the several readings is used to approve or disapprove the site quality. It is desirable to select the reading locations at random to maintain the objectivity of the statistical approach.

The following RANDOM SAMPLE CARDS provide this objectivity in selection. They were produced by a random computer plot, different for each card.


- 1. Select a card from a shuffled deck. Visualize the card as a map of the project to be tested.
- 2. Use the average of five readings to pass or fail. Select High, Low, Even, or Odd. This will select which five combinations out of the ten spots on the card are to be used.
- 3. Identify the exact reading site by location data using the card as the guide. Then go to the five locations and take a nuclear tester reading at each site, using backscatter or transmission as required.
- 4. Average the five readings and pass or fail the site.
- 5. Should differences in soil characteristics be suspected, take a shovelful of soil from each site and prepare a composite Proctor for optimum value preparation.

NONBIASED SAMPLE CARD #1


NONBIASED SAMPLE CARD #2



Non-Biased Random Sample Cards 1 and 2



Non-Biased Random Sample Cards 3, 4, 5, and 6



Non-Biased Random Sample Cards 7, 8, 9, and 10

J-3/(J-4 blank)

APPENDIX K

STANDARD INTERNATIONAL (SI)

CONVERSION CHART

Radiological Units

bequerel	2.703 x 10 ⁻¹¹	curies
curies	3.700 x 10 ¹⁰	dis/sec
curies	2.220 x 10 ¹²	dis/min
curies	10 ³	millicuries
curies	10 ⁶	microcuries
curies	10 ¹²	picocuries
curies	10 ⁻³	kilocuries
curies	3.700x 1010	bequerel
dis/min	4.505 x 10 ⁻¹⁰	millicuries
dis/min	4.505 x 10 ⁻⁷	microcuries
dis/sec	2.703 x 10 ⁻⁸	millicuries
dis/sec	2.703 x10- ⁵	microcuries
Multiply # of to obtain # of	by by	to obtain # of Divide # of
gray	100	rad
kilocuries	10 ³	curies
microcuries	3.700 x 10 ⁴	dis/sec
microcuries	2.220 x 10 ⁶	dis/min

APPENDIX K (CONT)

Multiply # of to obtain # of	by by	to obtain # of Divide # of
millicuries	3.700 x 10 ⁷	dis/sec
millicuries	2.220 x 10 ⁹	dis/min
R	2.58 x 10 ⁻⁴	C/kg of air
R	1	esu/cm ³ of air (s.t.p.)
R	2.082 x 10 ⁹	ion prs/cm ³ of air (s.t.p.)
R	1.610 x 10 ¹²	ion prs/g of air
R (33.7 eV/ion pr.)	7.02×10^4	meV/cm ³ of air (s.t.p.)
R (33.7 eV/ion pr.)	5.43 x 10 ⁷	meV/g of air
R (33.7 eV/ion pr.)	86.9	ergs/g of air
R (33.7 eV/ion pr.)	2.08 x 10 ⁻⁶	g-calg of air
R (33.7 eV/ion pr.)	x 98	ergs/g
rads	0.01	gray
rads	0.01	J/kg
rads	100	ergs/g of soft tissue
rads	8.071 x 10 ⁴	MeV/cm ³ or air
(s.t.p.) rads	6.242 x 10 ⁷	meV/g
rads	10 ⁻⁵	watt-sec/g

APPENDIX K (CONT)

Multiply # ofto obtain # of	by by	to obtain # of Divide # of
rads (33.7 ev/ion pr.)	2.39 x 109	ion prs/cm3 of air (s.t.p.)
rem	0.01	sievert
uCi/cm³ (Ci/mi)	2.22 x 10 ¹²	dpm/m ³
uCi/cm ³	2.22 x 10 ⁹	dpm/liter
dpm/m3	0.4505	pCi/m ³
sievert	100	rem

K-3/(K-4 blank)

MOISTURE & DENSITY TESTER FIELD DATA WORKSHEET						
				DATE: SITE:		
STANDARD COUNTS						
DENSITY COUNT	FS (ADD ZERO) HIGH SqRt + AVG SqRt - LOW		MOISTURE	AVG		Α
■	AVG =			√ AVG =		
	TOTAL StCt AVG StCt (÷10)			TOTAL St AVG StCt	<u>Ct</u> (÷10)	
FIELD COUNTS B/S	S DENSITY TRANS	. 2" TRANS. 4"	TRANS. 6" TRANS	5. 8"	B/S MOISTURE	р р
AVG ÷ 3 ÷ by AVG StCt				AVG ÷ 3		B C A
RESULTANT RATIO WET DEN PCF				H ₂ O PCF		C DW
DRY WEIGHT WET DEN PCF LESS H ₂ O DRY DEN PCF	(BS)	(2")	(4")	(6'')	(8'')	D DW E
PERCENT COMPACTION DRY DEN PCF *SOIL MAX DEN						E F
X 100 = % PERCENT MOISTURE H ₂ O PCF DRY DEN PCF						G DW E
X 100 = % MOISTURE						н

For use of this form, see TM 5-6635-386-12&P; the proponent agency is AMC.

DA FORM 5448-R, AUG 85

L-1/ (L-2 blank)

DA FORM 5448-R, AUG 85

^{*} THE MAXIMUM DENSITY OF SOIL REQUIRES A PROCTOR METHOD RUN UNDER LABORATORY CONDITIONS. THIS IS NOT SUPPLIED BY THE NUCLEAR TESTER. ALL % MOISTURE IS A PERCENT OF DRY WEIGHT, NOT WET WEIGHT.

APPENDIX M

MOISTURE & DENSITY TESTER (NUCLEAR METHOD) UTILIZATION LOG					
DATE	TIME PERIOD FROM TO	LOCATION OF USE	OPERATOR	REMARKS (include standard count for moisture and density)	

DA FORM 5449-R, AUG 85

For use of this form, see TM 5-6635-386-12&P; the proponent agency is AMC.

M-1/(M-2 blank)

APPENDIX N

STANDARD COUNT	DENSITY	MOISTURE	MOISTURE & DENSITY FIELD TEST SHEE	T FILE NO.
			GAUGE NO	PROJECT
TOTAL AVERAGES			MOISTURE & DENSITY DATA	
DATE				
TEST NO.				
TEST LOCATION				
PROBE DEPTH				
DENSITY COUNTS				
TOTAL				
AVERAGES				
DENSITY RATIO				
MOISTURE COUNTS				
TOTAL				
AVERAGES				
MOISTURE RATIO				
WET DENSITY p.c.f.				
WATER p.c.f.				
% MOISTURE				
DRY DENSITY p.c.f.				
% RELATIVE COMPACTION				

DA FORM 5450-R, AUG 85

For use of this form, see TM 5-6635-386-12&P; the proponent agency is AMC.

N-1/(N-2 blank)

INDEX

		Paragraph
	Α	
Asphalt Content Determination		1-31
Area Concept Testing		APPENDIX J
Authorized Stockage List (ASL)		
Safety		
Tester Requisition and Disposition		
Training Courses		4-39
	В	
Basic Issue Items List		APPENDIX C
Batteries		3-2
Charging		
General		
Replacement		3-5
Testing		3-4
	С	
Calibration Charts		1-28
Common Radiation Terms		
Completion of Test		1-27
Components of End Item and Basic Issue Items List		
Compute Percent Compaction		
Compute Percent Moisture		1-26
	D	
Density and Moisture Tester		
General Operation		1-1
Modes of Operation		
Principles of Operation		
Destruction to Prevent Enemy Use		
Determine Density and Moisture from Calibration Charts		
Determine Dry Weight		1-24
Diagnostic Questions		3-9
-		

INDEX-CONTINUED

		Paragraph
	E	
Emergency Procedures		
Damaged or Leaking Tester		. 4-31
Decontamination of Equipment and Area		. 4-34
Leaks, Loss or Tester Damage (Fire and Explosion)		. 4-29
Personnel Exposed to Excessive Radiation		
Personnel Exposed to Radiation		
Physical Loss of Tester		
Equipment Improvement Recommendations		
Equipment Publications		
Expendable/Durable Supplies and Materials List		. APPENDIX E
	F	
Facilities		. 5-8
Field Counts and Transmission		
Field Counts, Transmission and Moisture		
Fuses		. 3-10
	G	
General Operation		. 1-1
·	L	
	L	
Leak Test Procedure		. 4-26
Letter of Instruction		. APPENDIX G
	M	
Maintenance		
Tester Maintenance (Exterior)		. 3-8
Tester Maintenance (General)		. 4-35
Tester Repair and Disposal		
Maintenance Allocation Chart		. APPENDIX B
Maintenance Concept		
Major Sub-Assemblies Parts, MC-1		
Manufacturer's Field Campaigns and Modifications		
Manufacturer's Safety Defects		
Measure or Take a Test, General		
Milstrip Requisition For CCE (NON-NSN), Sample Format		
Milstrip Requisition For CCE (NSN), Sample Format		APPENDIX I

INDEX-CONTINUED

		Paragraph
M-CO	NTINUED	
Moisture Measurement Errors		. 1-29
Moisture Reflection Errors		. 1-30
	••	
	N	
Nuclear Reading On Site, Take a		1-20
rtasioai rtoaanig on ollo, railo a illiilliilliilliilliilliilliilliillii		20
	0	
Operator Maintenance, General		. 3-1
	Р	
	r	
Panel Controls		.1-4
Personnel and Training		
Personnel Dosimetry		
Film Badges or TLDs		
Medical Records		
Prescribed Load List (PLL)		APPENDIX D
	R	
Radiacmeters		
Radiation Practices		
References		
Repair Parts and SupportResponsibility		. 5-5
Alternate Local Radiation Protection Officer (ALRPO)		4-11
Commander of Major Command		
Local Radiation Protection Officer (LRPO)		
NICP Item Manager		
Operator		
Program Manager		
Radiation Control Officer		. 4-8
Tester Disposal		
Using Unit Commander		
Roof Moisture Testing		
Plot Roof Moisture Profile		
Procedure		
Radiation SafetyRelative Measurement		
Testing		
1 000119		02

INDEX-CONTINUED

S

	Paragraph
Safety Procedures, Purpose	4-1
Seating Quality Experiment	
Depth of Readingl-16	
Drill Transmission Hole	
Seating Tester	
Surface of Roughness Error	
Transmission Roughness Error	
Shipment	
Shipping Case Maintenance	
Shutter Maintenance	
Site Preparation	
Source Actuation .	
Standard Count,	
Calculator Evaluation Using Standard Count	1-9
How to Take	
Manual Evaluation	
Meaning of Poor Results	
Standard International Conversion Chart	
Storage	
Administration	4-19
Occupied Areas	4-18
Unoccupied Areas	4-17
Support Organizations	5-4
Surveys and Tests	
Minimum Requirements for Scheduled Surveys	4-24
Radiation Leak (Wipe) Test	4-25
Radiation Surveys	4-23
Т	
Transportation of Testers by Using Units	
Air Drop	4-21
Air Transportation	
General	
Conordi	± 20

Index 4

By Order of the Secretary of the Army:

GORDON R. SULLIVAN General, United States Army Chief of Staff

Official:

PATRICIA P. HICKERSON Brigadier General, United States Army The Adjutant General

Distribution:

To be distributed in accordance with DA Form 12-25-E (Block 2560). requirements for TM5-6635-386-12.

Unit maintenance

* U.S. GOVERNMENT PRINTING OFFICE: 1994 O - 300-421 (03142)

RECOMMENDED CHANGES TO EQUIPMENT TECHNICAL PUBLICATIONS				
752	Something	WRONG WITH THIS PUBLICATION?		
THEN JOT DOWN THE DOPE ABOUT IT ON THIS FORM, CAREFULLY TEAR IT OUT, FOLD IT AND DROP IT IN THE MAIL! FROM: (PRINT YOUR UNIT'S COMPLETE ADDRESS) DATE SENT				
PUBLICATION NUMBER	PUBLICATION DATE	PUBLICATION TITLE		
BE EXACTPIN-POINT WHERE IT IS PAGE PARA- FIGURE TABLE AND NO. GRAPH NO. NO.	THIS SPACE TELL WHAT II ID WHAT SHOULD BE DON	B WRONG E ABOUT IT:		
PRINTED NAME, GRADE OR TITLE, AND TELEPHONE	NUMBER SIGN H	ERE:		

DA 150E7. 2028-2

PREVIOUS EDITIONS
• ARE OBSOLETE.

P.S.—IF YOUR OUTFIT WANTS TO KNOW ABOUT YOUR RECOMMENDATION MAKE A CARBON COPY OF THIS AND GIVE IT TO YOUR HEADQUARTERS.

THE METRIC SYSTEM AND EQUIVALENTS

LINEAR MEASURE

- 1 Centimeter = 10 Millimeters = 0.01 Meters = 0.3937 Inches
- 1 Meter = 100 Centimeters = 1.000 Millimeters = 39.37 Inches
- 1 Kilometer = 1.000 Meters = 0.621 Miles

SQUARE MEASURE

- 1 Sq Centimeter = 100 Sq Millimeters = 0.155 Sq Inches
- 1 Sq Meter = 10,000 Sq Centimeters = 10.76 Sq Feet
- 1 Sq Kilometer = 1,000,000 Sq Meters = 0,386 Sq Miles

CUBIC MEASURE

1 Cu Centimeter = 1.000 Cu Millimeters = 0.06 Cu Inches

1 Cu Meter = 1.000.000 Cu Centimeters = 35.31 Cu Feet

LIQUID MEASURE

1 Milliliter = 0.001 Liters = 0.0338 Fluid Ounces 1 Liter = 1.000 Milliters = 33.82 Fluid Ounces

TEMPERATURE

5/9 (°+ -32) = °C

212° Fahrenheit is equivalent to 100° Celsius

90° Fahrenheit is equivalent to 32.2° Celsius

32° Fahrenheit is equivalent to 0° Celsius

 $9/5 \text{ C}^{\circ} + 32 = \text{F}^{\circ}$

WEIGHTS

- 1 Gram = 0.001 Kilograms = 1,000 Milligrams = 0.035 Ounces
- 1 Kilogram = 1.000 Grams = 2.2 l b.

I Metric Ton = 1.000 Kilograms = 1 Megagram = _

1.1 Short Tons

APPROXIMATE CONVERSION FACTORS			0-3-00
TO CHANGE	то	MULTIPLY BY	CENTIME
Inches	Centimeters	2.540	
Feet	Meters	0.305	NCHES
Yards	Meters	0.914	1 99 🕮 🖼 1
Miles	Kilometers	1 609	\ S -38" (∰)
Square Inches	Square Centimeters	6.451	1 28
Square Feet	Square Meters	0.093	- <u></u>
Square Yards	Square Meters	0.836	! ~ - ∃
Square Miles	Square Kilometers	2.590	1 3
Acres	Square Hectometers	0.405	-] [
Cubic Feet	Cubic Meters	0.028	1 -3
Cubic Yards	Cubic Meters	0.765	1 1
Fluid Ounces	Milliliters	29.573	 - 3 -
Pints	Liters	0.473	│ -}
Ouarts	Liters	0.946	1 - II 1
Gallons	Laters	3.785	N
Ounces	Grams	28.349	!—— E
Pounds	Kilograms	0.454	} 1
Short Tons	Metric Tons	0.907	-
	Newton-Meters	1.356	} _#
Pound-Feet		6.895	1 4
Pounds Per Square Inch	Kilopascals	•	├
Miles Per Gallon	Kilometers Per Liter	0.425 1.609	 ≢ ` (
Miles Per Hour	Kilometers Per Hour		ω_====
TO CHANGE	TO	MULTIPLYBY	
Centimeters	Inches	0.394	
Meters	Feet	3.280	1 _\$=- \
Meters	Yards	1.094	
Kilometers	Miles	0.621	
Square Centimeters	Square Inches	0.155	│ <u></u> ★
Square Meters	Square Feet	10.764	
Square Meters	Square Yards	1.196	
Square Kilometers	Square Miles	0.386	
Square Hectometers	Acres	2.471	3
Cubic Meters	Cubic Feet	35.315	_ = =
Cubic Meters	Cubic Yards	1.308	1 _3 = {
Milliliters	Fluid Ounces	0.034	1 T F
Liters	Pints	2.113	. <u>-</u> 1€− ≅
Liters	Quarts	1.057	
Liters	Gallons	0.264	(∃E
Grams	Ounces	0.035	
Kilograms	Pounds	2.205	1 TE ~ [
Metric Tons	Short Tons	1.102	1
Newton-Meters	Pound-Feet	0.738	_ ₹ ~ 1
Kilopascals	Pounds Per Square Inch	0.145	
Kilometers Per Liter	Miles Per Gallon	2.354	"I E-
Kilometers Per Hour	Miles Per Hour	0.621	
			0-3-5

PIN: 058285-000

This fine document...

Was brought to you by me:



<u>Liberated Manuals -- free army and government manuals</u>

Why do I do it? I am tired of sleazy CD-ROM sellers, who take publicly available information, slap "watermarks" and other junk on it, and sell it. Those masters of search engine manipulation make sure that their sites that sell free information, come up first in search engines. They did not create it... They did not even scan it... Why should they get your money? Why are not letting you give those free manuals to your friends?

I am setting this document FREE. This document was made by the US Government and is NOT protected by Copyright. Feel free to share, republish, sell and so on.

I am not asking you for donations, fees or handouts. If you can, please provide a link to liberatedmanuals.com, so that free manuals come up first in search engines:

Free Military and Government Manuals

- SincerelyIgor Chudovhttp://igor.chudov.com/
- Chicago Machinery Movers